

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021723\
 Data File : VX034155.D
 Acq On : 17 Feb 2023 13:52
 Operator : JC/MD
 Sample : VX0217WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0217WBL01

Quant Time: Feb 17 14:18:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	399233	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	671044	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	578571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	250598	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	254490	47.332	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.660%
35) Dibromofluoromethane	5.385	113	212489	48.634	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.260%
50) Toluene-d8	8.647	98	804696	51.298	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.600%
62) 4-Bromofluorobenzene	11.079	95	284086	47.346	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021723\
Data File : VX034155.D
Acq On : 17 Feb 2023 13:52
Operator : JC/MD
Sample : VX0217WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0217WBL01

Quant Time: Feb 17 14:18:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

