

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021723\
 Data File : VX034161.D
 Acq On : 17 Feb 2023 16:18
 Operator : JC/MD
 Sample : 01559-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10-DUP-20230213

Quant Time: Feb 17 16:53:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	373753	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	630793	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	550244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	238444	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	225120	44.724	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.440%
35) Dibromofluoromethane	5.385	113	198443	48.317	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.640%
50) Toluene-d8	8.647	98	754289	51.153	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.300%
62) 4-Bromofluorobenzene	11.079	95	260808	46.240	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.480%
Target Compounds						
16) Acetone	2.374	43	33206	12.788	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021723\
Data File : VX034161.D
Acq On : 17 Feb 2023 16:18
Operator : JC/MD
Sample : 01559-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10-DUP-20230213

Quant Time: Feb 17 16:53:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

