

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX021723\  
 Data File : VX034163.D  
 Acq On : 17 Feb 2023 17:04  
 Operator : JC/MD  
 Sample : 01559-05  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BP-VPB-RW10-EB-20230214

Quant Time: Feb 20 00:35:29 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X021723W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 17 12:33:35 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 362437   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 603211   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 519690   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 221610   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 215539   | 44.157   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =      | 88.320%  |
| 35) Dibromofluoromethane    | 5.385  | 113   | 188400   | 47.970   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 95.940%  |
| 50) Toluene-d8              | 8.647  | 98    | 716883   | 50.839   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =      | 101.680% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 241263   | 44.730   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 89.460%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 11) Tert butyl alcohol      | 2.965  | 59    | 10199    | 9.342    | ug/l # | 81       |
| 16) Acetone                 | 2.386  | 43    | 19895    | 7.901    | ug/l   | 95       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX021723\  
Data File : VX034163.D  
Acq On : 17 Feb 2023 17:04  
Operator : JC/MD  
Sample : 01559-05  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BP-VPB-RW10-EB-20230214

Quant Time: Feb 20 00:35:29 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X021723W.M  
Quant Title : SW846 8260  
QLast Update : Fri Feb 17 12:33:35 2023  
Response via : Initial Calibration

