

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021723\
 Data File : VX034166.D
 Acq On : 17 Feb 2023 18:37
 Operator : JC/MD
 Sample : 01598-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Feb 20 00:36:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	384892	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	646356	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	556373	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	236662	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	226308	43.659	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.320%
35) Dibromofluoromethane	5.385	113	202398	48.094	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.180%
50) Toluene-d8	8.647	98	767735	50.811	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.620%
62) 4-Bromofluorobenzene	11.079	95	258265	44.686	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021723\
Data File : VX034166.D
Acq On : 17 Feb 2023 18:37
Operator : JC/MD
Sample : 01598-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Feb 20 00:36:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

