

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021723\
 Data File : VX034167.D
 Acq On : 17 Feb 2023 19:00
 Operator : JC/MD
 Sample : 01598-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Feb 20 00:36:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	362908	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	604441	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	522018	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	216764	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	214401	43.867	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.740%
35) Dibromofluoromethane	5.385	113	190557	48.420	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.840%
50) Toluene-d8	8.647	98	722973	51.167	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.340%
62) 4-Bromofluorobenzene	11.079	95	239733	44.356	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021723\
Data File : VX034167.D
Acq On : 17 Feb 2023 19:00
Operator : JC/MD
Sample : 01598-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Feb 20 00:36:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

