

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
 Data File : VX040291.D
 Acq On : 17 Feb 2024 10:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 19 03:49:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	78708	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129594	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128730	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	69742	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	55047	47.011	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.020%
35) Dibromofluoromethane	5.379	113	44089	48.716	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.440%
50) Toluene-d8	8.647	98	163571	48.317	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.640%
62) 4-Bromofluorobenzene	11.079	95	67524	48.906	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.820%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	46945	52.802	ug/l	96
3) Chloromethane	1.294	50	43299	47.228	ug/l	99
4) Vinyl Chloride	1.374	62	47957	49.498	ug/l	99
5) Bromomethane	1.599	94	33651	61.549	ug/l	99
6) Chloroethane	1.684	64	29768	48.717	ug/l	98
7) Trichlorofluoromethane	1.886	101	78918	50.721	ug/l	100
8) Diethyl Ether	2.130	74	29379	50.265	ug/l	91
9) 1,1,2-Trichlorotrifluo...	2.325	101	44980	53.894	ug/l	98
10) Methyl Iodide	2.453	142	55100	59.466	ug/l	92
11) Tert butyl alcohol	2.953	59	52514	230.242	ug/l #	91
12) 1,1-Dichloroethene	2.318	96	41981	52.243	ug/l	84
13) Acrolein	2.233	56	51009	242.298	ug/l	99
14) Allyl chloride	2.660	41	71154	51.024	ug/l	96
15) Acrylonitrile	3.056	53	130894	256.490	ug/l	98
16) Acetone	2.373	43	132938	276.427	ug/l	93
17) Carbon Disulfide	2.507	76	99596	46.946	ug/l	99
18) Methyl Acetate	2.696	43	57637	44.707	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	151652	53.568	ug/l	98
20) Methylene Chloride	2.788	84	48712	54.163	ug/l	94
21) trans-1,2-Dichloroethene	3.087	96	45080	49.257	ug/l	93
22) Diisopropyl ether	3.757	45	155830	55.099	ug/l	89
23) Vinyl Acetate	3.715	43	702204	278.519	ug/l	98
24) 1,1-Dichloroethane	3.605	63	87581	53.162	ug/l	100
25) 2-Butanone	4.550	43	190588	253.530	ug/l	97
26) 2,2-Dichloropropane	4.471	77	70783	55.814	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	55438	53.099	ug/l	94
28) Bromochloromethane	4.897	49	39416	56.353	ug/l	92
29) Tetrahydrofuran	4.995	42	121179	247.429	ug/l	96
30) Chloroform	5.092	83	95269	53.487	ug/l	95
31) Cyclohexane	5.464	56	69916	50.059	ug/l	96
32) 1,1,1-Trichloroethane	5.379	97	80262	53.564	ug/l	99
36) 1,1-Dichloropropene	5.690	75	65667	51.797	ug/l	97
37) Ethyl Acetate	4.708	43	73357	52.002	ug/l	98
38) Carbon Tetrachloride	5.672	117	65944	54.707	ug/l	93
39) Methylcyclohexane	7.379	83	76561	51.965	ug/l	96
40) Benzene	6.031	78	192087	54.181	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
 Data File : VX040291.D
 Acq On : 17 Feb 2024 10:55
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 19 03:49:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.910	41	40428	54.061	ug/l	93
42) 1,2-Dichloroethane	6.080	62	76096	54.914	ug/l	100
43) Isopropyl Acetate	6.336	43	118519	53.821	ug/l	98
44) Trichloroethene	7.123	130	52053	53.228	ug/l	98
45) 1,2-Dichloropropane	7.427	63	52862	57.286	ug/l	98
46) Dibromomethane	7.580	93	39012	57.404	ug/l	96
47) Bromodichloromethane	7.818	83	73303	55.978	ug/l	95
48) Methyl methacrylate	7.689	41	59718	52.999	ug/l	91
49) 1,4-Dioxane	7.653	88	22279	908.884	ug/l #	92
51) 4-Methyl-2-Pentanone	8.567	43	401651	267.151	ug/l	99
52) Toluene	8.714	92	126471	54.462	ug/l	95
53) t-1,3-Dichloropropene	8.976	75	83101	53.249	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	85704	58.398	ug/l	94
55) 1,1,2-Trichloroethane	9.147	97	55691	56.228	ug/l	98
56) Ethyl methacrylate	9.116	69	83739	56.476	ug/l	96
57) 1,3-Dichloropropane	9.305	76	89509	55.368	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	185128	263.931	ug/l	98
59) 2-Hexanone	9.427	43	322510	270.558	ug/l	98
60) Dibromochloromethane	9.518	129	58484	57.957	ug/l	99
61) 1,2-Dibromoethane	9.610	107	59676	56.684	ug/l	98
64) Tetrachloroethene	9.268	164	45041	49.929	ug/l	96
65) Chlorobenzene	10.079	112	141791	52.776	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	52212	56.112	ug/l	98
67) Ethyl Benzene	10.189	91	260841	54.513	ug/l	100
68) m/p-Xylenes	10.299	106	195614	108.245	ug/l	97
69) o-Xylene	10.640	106	97517	55.530	ug/l	98
70) Styrene	10.652	104	165087	56.543	ug/l	99
71) Bromoform	10.799	173	44734	51.149	ug/l #	99
73) Isopropylbenzene	10.963	105	266802	54.315	ug/l	99
74) N-amyl acetate	10.841	43	117570	55.838	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.213	83	93703	53.096	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	100551	71.196	ug/l	90
77) Bromobenzene	11.195	156	63542	54.769	ug/l	96
78) n-propylbenzene	11.305	91	327163	55.013	ug/l	99
79) 2-Chlorotoluene	11.360	91	191771	53.401	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	230274	55.683	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	25156	50.017	ug/l	95
82) 4-Chlorotoluene	11.451	91	228700	53.750	ug/l	99
83) tert-Butylbenzene	11.713	119	221616	54.661	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	230275	55.039	ug/l	100
85) sec-Butylbenzene	11.890	105	294716	56.562	ug/l	100
86) p-Isopropyltoluene	12.006	119	239925	56.173	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	127281	55.369	ug/l	98
88) 1,4-Dichlorobenzene	12.042	146	130138	53.462	ug/l	98
89) n-Butylbenzene	12.329	91	231524	56.155	ug/l	98
90) Hexachloroethane	12.536	117	39211	58.201	ug/l	91
91) 1,2-Dichlorobenzene	12.335	146	123204	55.364	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21826	53.585	ug/l	90
93) 1,2,4-Trichlorobenzene	13.585	180	85267	55.967	ug/l	95
94) Hexachlorobutadiene	13.725	225	34134	56.137	ug/l	100
95) Naphthalene	13.774	128	264224	54.911	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	81374	56.034	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
Data File : VX040291.D
Acq On : 17 Feb 2024 10:55
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Feb 19 03:49:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:168 Resp: 78708
Ion Ratio Lower Upper
168 100
99 56.9 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 52.802 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55


Tgt Ion: 85 Resp: 46945
Ion Ratio Lower Upper
85 100
87 31.0 16.7 50.1








#3
Chloromethane
Concen: 47.228 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 50 Resp: 43299
Ion Ratio Lower Upper
50 100
52 32.5 25.5 38.3



#4
Vinyl Chloride
Concen: 49.498 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion: 62 Resp: 47957
Ion Ratio Lower Upper
62 100
64 32.0 25.4 38.0









#5
Bromomethane
Concen: 61.549 ug/l
RT: 1.599 min Scan# 84
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 33651
Ion Ratio Lower Upper
94 100
96 94.1 74.6 111.8



#6
Chloroethane
Concen: 48.717 ug/l
RT: 1.684 min Scan# 98
Delta R.T. 0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion: 64 Resp: 29768
Ion Ratio Lower Upper
64 100
66 30.2 25.1 37.7









#7
Trichlorofluoromethane
Concen: 50.721 ug/l
RT: 1.886 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 78918
Ion Ratio Lower Upper
101 100
103 64.1 51.4 77.2



#8
Diethyl Ether
Concen: 50.265 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion: 74 Resp: 29379
Ion Ratio Lower Upper
74 100
45 96.7 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.894 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 44980
Ion Ratio Lower Upper
101 100
85 46.0 37.0 55.6
151 76.4 58.6 87.8



#10
Methyl Iodide
Concen: 59.466 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion:142 Resp: 55100
Ion Ratio Lower Upper
142 100
127 44.7 41.6 62.4
141 14.1 11.6 17.4





#11

Tert butyl alcohol

Concen: 230.242 ug/l

RT: 2.953 min Scan# 308

Delta R.T. -0.012 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 52514

Ion Ratio Lower Upper

59 100

57 10.7 6.0 9.0#



#12

1,1-Dichloroethene

Concen: 52.243 ug/l

RT: 2.318 min Scan# 202

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55




Tgt Ion: 96 Resp: 41981


Ion Ratio Lower Upper

96 100

61 152.1 144.6 216.8

98 59.7 50.4 75.6





#13

Acrolein

Concen: 242.298 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 51009

Ion Ratio Lower Upper

56 100

55 70.9 57.1 85.7



#14

Allyl chloride

Concen: 51.024 ug/l

RT: 2.660 min Scan# 258

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55




Tgt Ion: 41 Resp: 71154


Ion Ratio Lower Upper

41 100

39 70.9 55.8 83.6


76 35.3 24.3 36.5








#15
Acrylonitrile
Concen: 256.490 ug/l
RT: 3.056 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:	53	Resp:	130894
Ion	Ratio	Lower	Upper
53	100		
52	84.3	66.0	99.0
51	36.7	29.8	44.6



#16
Acetone
Concen: 276.427 ug/l
RT: 2.373 min Scan# 211
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55


Tgt Ion:	43	Resp:	132938
Ion	Ratio	Lower	Upper
43	100		
58	31.0	21.9	32.9








#17
Carbon Disulfide
Concen: 46.946 ug/l
RT: 2.507 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 76 Resp: 99596
Ion Ratio Lower Upper
76 100
78 9.2 7.0 10.6



#18
Methyl Acetate
Concen: 44.707 ug/l
RT: 2.696 min Scan# 264
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion: 43 Resp: 57637
Ion Ratio Lower Upper
43 100
74 23.2 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 53.568 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 151652
Ion Ratio Lower Upper
73 100
57 22.8 17.3 25.9



#20
Methylene Chloride
Concen: 54.163 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55


Tgt Ion: 84 Resp: 48712
Ion Ratio Lower Upper
84 100
49 122.9 103.6 155.4
51 35.9 33.8 50.6
86 65.2 49.4 74.0








#21
trans-1,2-Dichloroethene
Concen: 49.257 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 45080
Ion Ratio Lower Upper
96 100
61 139.3 119.2 178.8
98 63.6 48.7 73.1




#22
Diisopropyl ether
Concen: 55.099 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55



Tgt Ion: 45 Resp: 155830
Ion Ratio Lower Upper
45 100
43 74.2 69.7 104.5
87 28.7 20.8 31.2
59 11.7 9.0 13.4









#23
 Vinyl Acetate
 Concen: 278.519 ug/l
 RT: 3.715 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX040291.D
 Acq: 17 Feb 2024 10:55




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 702204
 Ion Ratio Lower Upper
 43 100
 86 10.4 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 53.162 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX040291.D
 Acq: 17 Feb 2024 10:55

Tgt Ion: 63 Resp: 87581
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.3 9.8
 100 4.2 2.1 6.5









#25
2-Butanone
Concen: 253.530 ug/l
RT: 4.550 min Scan# 50
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 190588
Ion Ratio Lower Upper
43 100
72 25.8 19.4 29.2



#26
2,2-Dichloropropane
Concen: 55.814 ug/l
RT: 4.471 min Scan# 555
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion: 77 Resp: 70783
Ion Ratio Lower Upper
77 100
97 22.6 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 53.099 ug/l

RT: 4.489 min Scan# 558

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 55438

Ion Ratio Lower Upper

96 100

61 145.4 0.0 312.2

98 64.6 0.0 129.0



#28

Bromochloromethane

Concen: 56.353 ug/l

RT: 4.897 min Scan# 625

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55




Tgt Ion: 49 Resp: 39416


Ion Ratio Lower Upper

49 100

129 2.2 0.0 3.8

130 78.8 57.5 86.3





#29

Tetrahydrofuran

Concen: 247.429 ug/l

RT: 4.995 min Scan# 64

Delta R.T. -0.012 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 121179

Ion Ratio Lower Upper

42 100

72 45.5 34.2 51.4

71 42.6 32.2 48.2



#30

Chloroform

Concen: 53.487 ug/l

RT: 5.092 min Scan# 657

Delta R.T. -0.006 min

Lab File: VX040291.D




Acq: 17 Feb 2024 10:55


Tgt Ion: 83 Resp: 95269

Ion Ratio Lower Upper

83 100


85 66.4 50.3 75.5








#31
Cyclohexane
Concen: 50.059 ug/l
RT: 5.464 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 69916
Ion Ratio Lower Upper
56 100
69 31.4 23.4 35.2
84 85.4 65.9 98.9



#32
1,1,1-Trichloroethane
Concen: 53.564 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion: 97 Resp: 80262
Ion Ratio Lower Upper
97 100
99 64.4 51.1 76.7
61 46.3 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 47.011 ug/l

RT: 5.952 min Scan# 79

Delta R.T. -0.006 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 55047

Ion Ratio Lower Upper

65 100

67 51.1 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.006 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55




Tgt Ion: 114 Resp: 129594


Ion Ratio Lower Upper

114 100

63 19.0 0.0 45.0


88 16.7 0.0 36.6








#35
 Dibromofluoromethane
 Concen: 48.716 ug/l
 RT: 5.379 min Scan# 704
 Delta R.T. -0.006 min
 Lab File: VX040291.D
 Acq: 17 Feb 2024 10:55

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion:113 Resp: 44089
 Ion Ratio Lower Upper
 113 100
 111 101.2 82.9 124.3
 192 18.7 13.3 19.9



#36
 1,1-Dichloropropene
 Concen: 51.797 ug/l
 RT: 5.690 min Scan# 755
 Delta R.T. -0.006 min
 Lab File: VX040291.D
 Acq: 17 Feb 2024 10:55

Tgt Ion: 75 Resp: 65667
 Ion Ratio Lower Upper
 75 100
 110 36.4 17.0 50.9
 77 31.2 24.7 37.1





#37

Ethyl Acetate

Concen: 52.002 ug/l

RT: 4.708 min Scan# 594

Delta R.T. -0.012 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 43 Resp: 73357

Ion Ratio Lower Upper

43 100

61 14.3 10.6 16.0

70 10.5 8.3 12.5



#38

Carbon Tetrachloride

Concen: 54.707 ug/l

RT: 5.672 min Scan# 752

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55




Tgt Ion: 117 Resp: 65944


Ion Ratio Lower Upper

117 100

119 89.2 78.2 117.2


121 30.3 24.4 36.6








#39
Methylcyclohexane
Concen: 51.965 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 83 Resp: 76561
Ion Ratio Lower Upper
83 100
55 75.8 64.6 97.0
98 46.2 36.7 55.1



#40
Benzene
Concen: 54.181 ug/l
RT: 6.031 min Scan# 811
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55


Tgt Ion: 78 Resp: 192087
Ion Ratio Lower Upper
78 100
77 22.3 19.3 28.9








#41
Methacrylonitrile
Concen: 54.061 ug/l
RT: 4.910 min Scan# 61
Delta R.T. -0.018 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:	41	Resp:	40428
Ion	Ratio	Lower	Upper
41	100		
39	57.2	45.2	67.8
67	73.5	52.6	78.8
52	33.7	22.7	34.1



#42
1,2-Dichloroethane
Concen: 54.914 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion:	62	Resp:	76096
Ion	Ratio	Lower	Upper
62	100		
98	8.1	0.0	16.0









#43
Isopropyl Acetate
Concen: 53.821 ug/l
RT: 6.336 min Scan# 861
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 118519
Ion Ratio Lower Upper
43 100
61 19.2 15.2 22.8
87 13.6 9.8 14.6



#44
Trichloroethene
Concen: 53.228 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion:130 Resp: 52053
Ion Ratio Lower Upper
130 100
95 104.0 0.0 204.4





#45

1,2-Dichloropropane

Concen: 57.286 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 52862

Ion Ratio Lower Upper

63 100

65 31.4 24.3 36.5



#46

Dibromomethane

Concen: 57.404 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55




Tgt Ion: 93 Resp: 39012


Ion Ratio Lower Upper

93 100

95 82.0 66.6 100.0

174 93.2 69.4 104.0









#47
Bromodichloromethane
Concen: 55.978 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 73303
Ion Ratio Lower Upper
83 100
85 66.3 49.8 74.6
127 8.3 7.1 10.7



#48
Methyl methacrylate
Concen: 52.999 ug/l
RT: 7.689 min Scan# 1083
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion: 41 Resp: 59718
Ion Ratio Lower Upper
41 100
69 84.0 59.2 88.8
39 63.7 47.7 71.5





#49

1,4-Dioxane

Concen: 908.884 ug/l

RT: 7.653 min Scan# 1077

Delta R.T. -0.006 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 22279

Ion Ratio Lower Upper

88 100

43 35.0 0.0 0.0#

58 68.5 60.1 90.1



#50

Toluene-d8

Concen: 48.317 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.006 min

Lab File: VX040291.D




Acq: 17 Feb 2024 10:55


Tgt Ion: 98 Resp: 163571

Ion Ratio Lower Upper

98 100

100 64.5 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 267.151 ug/l
RT: 8.567 min Scan# 11
Delta R.T. -0.007 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 401651
Ion Ratio Lower Upper
43 100
58 38.4 30.4 45.6



#52
Toluene
Concen: 54.462 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion: 92 Resp: 126471
Ion Ratio Lower Upper
92 100
91 174.7 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 53.249 ug/l

RT: 8.976 min Scan# 1193

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 83101

Ion Ratio Lower Upper

75 100

77 32.2 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 58.398 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. -0.006 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55




Tgt Ion: 75 Resp: 85704


Ion Ratio Lower Upper

75 100

77 32.5 25.3 37.9

39 51.9 46.7 70.1









#55
1,1,2-Trichloroethane
Concen: 56.228 ug/l
RT: 9.147 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:	97	Resp:	55691
Ion	Ratio	Lower	Upper
97	100		
83	82.6	65.8	98.8
85	54.2	41.6	62.4
99	62.2	47.8	71.8



#56
Ethyl methacrylate
Concen: 56.476 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion:	69	Resp:	83739
Ion	Ratio	Lower	Upper
69	100		
41	68.0	56.8	85.2
39	42.6	32.0	48.0





#57

1,3-Dichloropropane

Concen: 55.368 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 89509

Ion Ratio Lower Upper

76 100

78 32.9 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 263.931 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX040291.D




Acq: 17 Feb 2024 10:55


Tgt Ion: 63 Resp: 185128

Ion Ratio Lower Upper

63 100

106 27.1 22.5 33.7









#59
2-Hexanone
Concen: 270.558 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 322510
Ion Ratio Lower Upper
43 100
58 53.9 26.2 78.5



#60
Dibromochloromethane
Concen: 57.957 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion:129 Resp: 58484
Ion Ratio Lower Upper
129 100
127 78.1 38.8 116.4









#61
1,2-Dibromoethane
Concen: 56.684 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 59676
Ion Ratio Lower Upper
107 100
109 91.9 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 48.906 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion: 95 Resp: 67524
Ion Ratio Lower Upper
95 100
174 77.9 0.0 135.6
176 74.7 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:117 Resp: 128730
Ion Ratio Lower Upper
117 100
82 54.0 46.2 69.4
119 31.9 25.5 38.3



#64
Tetrachloroethene
Concen: 49.929 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion:164 Resp: 45041
Ion Ratio Lower Upper
164 100
166 122.4 102.7 154.1
129 98.7 80.4 120.6
131 92.3 76.5 114.7









#65
Chlorobenzene
Concen: 52.776 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 141791
Ion Ratio Lower Upper
112 100
114 31.8 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 56.112 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion:131 Resp: 52212
Ion Ratio Lower Upper
131 100
133 95.1 47.1 141.4
119 67.0 32.2 96.6









#67
Ethyl Benzene
Concen: 54.513 ug/l
RT: 10.189 min Scan# 1494
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 260841
Ion Ratio Lower Upper
91 100
106 31.0 24.9 37.3



#68
m/p-Xylenes
Concen: 108.245 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion:106 Resp: 195614
Ion Ratio Lower Upper
106 100
91 212.2 165.5 248.3





#69

o-Xylene

Concen: 55.530 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:106 Resp: 97517

Ion Ratio Lower Upper

106 100

91 221.1 108.9 326.8



#70

Styrene

Concen: 56.543 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55




Tgt Ion:104 Resp: 165087


Ion Ratio Lower Upper

104 100

78 53.8 43.6 65.4

103 54.9 43.6 65.4





#71

Bromoform

Concen: 51.149 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX040291.D





Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:173	Resp:	44734
Ion Ratio	Lower	Upper
173	100	
175	48.8	24.1 72.4
254	0.0	0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.018 min Scan# 1793


Delta R.T. -0.006 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Tgt Ion:152	Resp:	69742
Ion Ratio	Lower	Upper
152	100	
115	89.0	43.3 129.9
150	175.9	0.0 345.0





#73

Isopropylbenzene

Concen: 54.315 ug/l

RT: 10.963 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 266802

Ion Ratio Lower Upper

105 100

120 26.3 12.9 38.6



#74

N-ethyl acetate

Concen: 55.838 ug/l

RT: 10.841 min Scan# 1600

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Tgt Ion: 43 Resp: 117570




Ion Ratio Lower Upper


43 100

70 40.9 32.9 49.3

55 24.2 19.3 28.9

61 23.8 18.8 28.2









#75
1,1,2,2-Tetrachloroethane
Concen: 53.096 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:	83	Resp:	93703
Ion	Ratio	Lower	Upper
83	100		
131	10.0	5.3	15.8
85	64.5	32.0	96.2



#76
1,2,3-Trichloropropane
Concen: 71.196 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion:	75	Resp:	100551
Ion	Ratio	Lower	Upper
75	100		
77	32.2	19.1	57.5





#77

Bromobenzene

Concen: 54.769 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.006 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:156 Resp: 63542

Ion Ratio Lower Upper

156 100

77 169.5 88.4 265.3

158 97.9 48.0 144.0



#78

n-propylbenzene

Concen: 55.013 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX040291.D




Acq: 17 Feb 2024 10:55


Tgt Ion: 91 Resp: 327163

Ion Ratio Lower Upper

91 100

120 22.2 11.5 34.4









#79
2-Chlorotoluene
Concen: 53.401 ug/l
RT: 11.360 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 191771
Ion Ratio Lower Upper
91 100
126 33.2 16.3 48.8



#80
1,3,5-Trimethylbenzene
Concen: 55.683 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion:105 Resp: 230274
Ion Ratio Lower Upper
105 100
120 47.0 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 50.017 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 25156
Ion Ratio Lower Upper
75 100
53 106.2 90.5 135.7
89 48.2 36.9 55.3



#82
4-Chlorotoluene
Concen: 53.750 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion: 91 Resp: 228700
Ion Ratio Lower Upper
91 100
126 28.7 14.2 42.6





#83

tert-Butylbenzene

Concen: 54.661 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:119 Resp: 221616

Ion Ratio Lower Upper

119 100

91 63.1 31.2 93.6

134 22.9 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 55.039 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX040291.D




Acq: 17 Feb 2024 10:55


Tgt Ion:105 Resp: 230275

Ion Ratio Lower Upper

105 100

120 42.6 21.2 63.6









#85
sec-Butylbenzene
Concen: 56.562 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 294716
Ion Ratio Lower Upper
105 100
134 19.4 9.8 29.4



#86
p-Isopropyltoluene
Concen: 56.173 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX040291.D
Acq: 17 Feb 2024 10:55

Tgt Ion:119 Resp: 239925
Ion Ratio Lower Upper
119 100
134 26.5 12.8 38.3
91 26.4 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 55.369 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 127281

Ion Ratio Lower Upper

146 100

111 42.4 22.1 66.1

148 64.2 31.8 95.3



#88

1,4-Dichlorobenzene

Concen: 53.462 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55




Tgt Ion:146 Resp: 130138


Ion Ratio Lower Upper

146 100

111 40.6 21.3 63.7

148 64.5 31.8 95.4





#89

n-Butylbenzene

Concen: 56.155 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.006 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 231524

Ion Ratio Lower Upper

91 100

92 52.7 27.4 82.0

134 25.4 12.6 37.8



#90

Hexachloroethane

Concen: 58.201 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX040291.D




Acq: 17 Feb 2024 10:55


Tgt Ion: 117 Resp: 39211

Ion Ratio Lower Upper

117 100

201 73.0 32.9 98.6





#91

1,2-Dichlorobenzene

Concen: 55.364 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 123204

Ion Ratio Lower Upper

146 100

111 43.6 22.4 67.0

148 63.0 31.6 95.0



#92

1,2-Dibromo-3-Chloropropane

Concen: 53.585 ug/l

RT: 12.939 min Scan# 1944

Delta R.T. -0.006 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55




Tgt Ion: 75 Resp: 21826


Ion Ratio Lower Upper

75 100

155 80.0 36.1 108.5

157 100.1 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 55.967 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 85267

Ion Ratio Lower Upper

180 100

182 92.8 48.3 144.8

145 30.8 18.0 54.0



#94

Hexachlorobutadiene

Concen: 56.137 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55




Tgt Ion:225 Resp: 34134


Ion Ratio Lower Upper

225 100

223 62.3 31.1 93.2

227 64.7 32.5 97.5





#95

Naphthalene

Concen: 54.911 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min


Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Instrument :



MSVOA_X

ClientSampleId :



Tgt Ion:128 Resp: 264224

Ion	Ratio	Lower	Upper
128	100		
127	12.8	10.3	15.5
129	10.8	8.8	13.2



#96

1,2,3-Trichlorobenzene

Concen: 56.034 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX040291.D

Acq: 17 Feb 2024 10:55

Tgt Ion:180 Resp: 81374

Ion	Ratio	Lower	Upper
180	100		
182	94.0	46.9	140.6
145	33.3	17.8	53.5

