

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
 Data File : VX040316.D
 Acq On : 17 Feb 2024 20:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 19 04:02:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	68904	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	116060	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61912	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51399	50.141	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 100.280%			
35) Dibromofluoromethane	5.379	113	41533	51.244	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.480%			
50) Toluene-d8	8.647	98	151910	50.105	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.200%			
62) 4-Bromofluorobenzene	11.079	95	63417	51.287	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 102.580%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.167	85	40304	51.782	ug/l	99
3) Chloromethane	1.295	50	40314	50.228	ug/l	99
4) Vinyl Chloride	1.374	62	43599	51.403	ug/l	100
5) Bromomethane	1.599	94	28458	59.457	ug/l	99
6) Chloroethane	1.679	64	27145	50.745	ug/l	100
7) Trichlorofluoromethane	1.886	101	70555	51.798	ug/l	100
8) Diethyl Ether	2.130	74	26070	50.950	ug/l	90
9) 1,1,2-Trichlorotrifluo...	2.331	101	38730	53.008	ug/l	98
10) Methyl Iodide	2.453	142	47010	57.954	ug/l	92
11) Tert butyl alcohol	2.959	59	52277	261.815	ug/l	# 91
12) 1,1-Dichloroethene	2.319	96	37477	53.273	ug/l	88
13) Acrolein	2.233	56	45541	247.104	ug/l	99
14) Allyl chloride	2.660	41	65075	53.304	ug/l	96
15) Acrylonitrile	3.063	53	124962	279.707	ug/l	100
16) Acetone	2.374	43	101555	241.216	ug/l	94
17) Carbon Disulfide	2.514	76	87953	47.357	ug/l	99
18) Methyl Acetate	2.703	43	55984	49.604	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	136548	55.096	ug/l	99
20) Methylene Chloride	2.788	84	44305	56.337	ug/l	95
21) trans-1,2-Dichloroethene	3.093	96	41322	51.575	ug/l	89
22) Diisopropyl ether	3.758	45	138988	56.136	ug/l	98
23) Vinyl Acetate	3.721	43	623868	282.656	ug/l	98
24) 1,1-Dichloroethane	3.605	63	78118	54.165	ug/l	100
25) 2-Butanone	4.550	43	171196	260.137	ug/l	97
26) 2,2-Dichloropropane	4.471	77	51827	46.682	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	49216	53.847	ug/l	93
28) Bromochloromethane	4.898	49	32232	52.639	ug/l	93
29) Tetrahydrofuran	4.995	42	116381	271.444	ug/l	97
30) Chloroform	5.093	83	85428	54.786	ug/l	98
31) Cyclohexane	5.471	56	61143	50.007	ug/l	94
32) 1,1,1-Trichloroethane	5.385	97	70062	53.409	ug/l	98
36) 1,1-Dichloropropene	5.696	75	58422	51.456	ug/l	98
37) Ethyl Acetate	4.709	43	67847	53.705	ug/l	99
38) Carbon Tetrachloride	5.678	117	56962	52.766	ug/l	99
39) Methylcyclohexane	7.379	83	64423	48.825	ug/l	94
40) Benzene	6.038	78	172490	54.327	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
 Data File : VX040316.D
 Acq On : 17 Feb 2024 20:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 19 04:02:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.910	41	35914	53.625	ug/l #	91
42) 1,2-Dichloroethane	6.086	62	67070	54.044	ug/l	99
43) Isopropyl Acetate	6.336	43	107053	54.283	ug/l	98
44) Trichloroethene	7.123	130	48317	55.169	ug/l	99
45) 1,2-Dichloropropane	7.428	63	46687	56.494	ug/l	93
46) Dibromomethane	7.580	93	34552	56.770	ug/l	95
47) Bromodichloromethane	7.818	83	67454	57.519	ug/l	99
48) Methyl methacrylate	7.690	41	54459	53.968	ug/l	92
49) 1,4-Dioxane	7.659	88	23986	1092.629	ug/l #	89
51) 4-Methyl-2-Pentanone	8.574	43	380386	282.511	ug/l	100
52) Toluene	8.714	92	112607	54.146	ug/l	96
53) t-1,3-Dichloropropene	8.976	75	69796	50.013	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	73229	55.717	ug/l	91
55) 1,1,2-Trichloroethane	9.153	97	49008	55.251	ug/l	93
56) Ethyl methacrylate	9.116	69	74215	55.889	ug/l	95
57) 1,3-Dichloropropane	9.305	76	80598	55.669	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	160290	255.169	ug/l	98
59) 2-Hexanone	9.427	43	295681	276.977	ug/l	98
60) Dibromochloromethane	9.519	129	50559	55.946	ug/l	99
61) 1,2-Dibromoethane	9.610	107	51889	55.035	ug/l	99
64) Tetrachloroethene	9.275	164	41199	51.785	ug/l	97
65) Chlorobenzene	10.080	112	123778	52.240	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	45823	55.839	ug/l	99
67) Ethyl Benzene	10.189	91	226888	53.766	ug/l	98
68) m/p-Xylenes	10.299	106	170463	106.957	ug/l	99
69) o-Xylene	10.640	106	83861	54.147	ug/l	98
70) Styrene	10.653	104	143903	55.886	ug/l	98
71) Bromoform	10.799	173	38053	49.396	ug/l #	99
73) Isopropylbenzene	10.964	105	230237	52.799	ug/l	100
74) N-amyl acetate	10.842	43	103683	55.471	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.214	83	85545	54.603	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	89519	71.401	ug/l	89
77) Bromobenzene	11.195	156	54947	53.350	ug/l	97
78) n-propylbenzene	11.305	91	278297	52.714	ug/l	98
79) 2-Chlorotoluene	11.366	91	162779	51.061	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	194488	52.977	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.018	75	20859	46.953	ug/l	96
82) 4-Chlorotoluene	11.451	91	193374	51.195	ug/l	99
83) tert-Butylbenzene	11.713	119	185994	51.676	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	197006	53.042	ug/l	100
85) sec-Butylbenzene	11.890	105	246364	53.262	ug/l	100
86) p-Isopropyltoluene	12.006	119	200722	52.938	ug/l	98
87) 1,3-Dichlorobenzene	11.970	146	106824	52.347	ug/l	98
88) 1,4-Dichlorobenzene	12.043	146	108592	50.253	ug/l	98
89) n-Butylbenzene	12.329	91	190072	51.931	ug/l	98
90) Hexachloroethane	12.536	117	32501	54.343	ug/l	92
91) 1,2-Dichlorobenzene	12.335	146	105871	53.592	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	19837	54.861	ug/l	94
93) 1,2,4-Trichlorobenzene	13.585	180	68042	50.309	ug/l	97
94) Hexachlorobutadiene	13.725	225	26481	49.059	ug/l	97
95) Naphthalene	13.774	128	224958	52.664	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	68120	52.840	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
Data File : VX040316.D
Acq On : 17 Feb 2024 20:28
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Feb 19 04:02:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Feb 19 04:02:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 68904
Ion Ratio Lower Upper
168 100
99 56.3 56.6 85.0#



#2
Dichlorodifluoromethane
Concen: 51.782 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 85 Resp: 40304
Ion Ratio Lower Upper
85 100
87 32.8 16.7 50.1









#3
Chloromethane
Concen: 50.228 ug/l
RT: 1.295 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 40314
Ion Ratio Lower Upper
50 100
52 31.4 25.5 38.3



#4
Vinyl Chloride
Concen: 51.403 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 62 Resp: 43599
Ion Ratio Lower Upper
62 100
64 31.9 25.4 38.0









#5
Bromomethane
Concen: 59.457 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 28458
Ion Ratio Lower Upper
94 100
96 92.1 74.6 111.8



#6
Chloroethane
Concen: 50.745 ug/l
RT: 1.679 min Scan# 97
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 64 Resp: 27145
Ion Ratio Lower Upper
64 100
66 31.3 25.1 37.7









#7
Trichlorofluoromethane
Concen: 51.798 ug/l
RT: 1.886 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 70555
Ion Ratio Lower Upper
101 100
103 64.2 51.4 77.2



#8
Diethyl Ether
Concen: 50.950 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 74 Resp: 26070
Ion Ratio Lower Upper
74 100
45 96.5 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.008 ug/l
RT: 2.331 min Scan# 203
Delta R.T. 0.006 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 38730
Ion Ratio Lower Upper
101 100
85 45.9 37.0 55.6
151 76.0 58.6 87.8



#10
Methyl Iodide
Concen: 57.954 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion:142 Resp: 47010
Ion Ratio Lower Upper
142 100
127 44.8 41.6 62.4
141 13.7 11.6 17.4





#11

Tert butyl alcohol

Concen: 261.815 ug/l

RT: 2.959 min Scan# 308

Delta R.T. -0.006 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 52277

Ion Ratio Lower Upper

59 100

57 10.7 6.0 9.0#



#12

1,1-Dichloroethene

Concen: 53.273 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28




Tgt Ion: 96 Resp: 37477


Ion Ratio Lower Upper

96 100

61 159.4 144.6 216.8

98 61.5 50.4 75.6





#13

Acrolein

Concen: 247.104 ug/l

RT: 2.233 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 45541

Ion Ratio Lower Upper

56 100

55 72.4 57.1 85.7



#14

Allyl chloride

Concen: 53.304 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28




Tgt Ion: 41 Resp: 65075


Ion Ratio Lower Upper

41 100

39 67.1 55.8 83.6

76 33.6 24.3 36.5





#15

Acrylonitrile

Concen: 279.707 ug/l

RT: 3.063 min Scan# 31

Delta R.T. 0.000 min


Lab File: VX040316.D

Acq: 17 Feb 2024 20:28




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:	53	Resp:	124962
Ion	Ratio	Lower	Upper
53	100		
52	82.4	66.0	99.0
51	36.7	29.8	44.6



#16

Acetone

Concen: 241.216 ug/l




RT: 2.374 min Scan# 211


Delta R.T. -0.006 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28


Tgt Ion:	43	Resp:	101555
Ion	Ratio	Lower	Upper
43	100		
58	30.6	21.9	32.9








#17
Carbon Disulfide
Concen: 47.357 ug/l
RT: 2.514 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 76 Resp: 87953
Ion Ratio Lower Upper
76 100
78 9.3 7.0 10.6



#18
Methyl Acetate
Concen: 49.604 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 43 Resp: 55984
Ion Ratio Lower Upper
43 100
74 23.8 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 55.096 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 136548
Ion Ratio Lower Upper
73 100
57 21.1 17.3 25.9



#20
Methylene Chloride
Concen: 56.337 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 84 Resp: 44305
Ion Ratio Lower Upper
84 100
49 125.4 103.6 155.4
51 37.0 33.8 50.6
86 66.9 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 51.575 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 41322
Ion Ratio Lower Upper
96 100
61 131.5 119.2 178.8
98 57.2 48.7 73.1



#22
Diisopropyl ether
Concen: 56.136 ug/l
RT: 3.758 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 45 Resp: 138988
Ion Ratio Lower Upper
45 100
43 85.4 69.7 104.5
87 28.2 20.8 31.2
59 11.7 9.0 13.4









#23
 Vinyl Acetate
 Concen: 282.656 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX040316.D
 Acq: 17 Feb 2024 20:28




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 623868
 Ion Ratio Lower Upper
 43 100
 86 10.5 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 54.165 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX040316.D
 Acq: 17 Feb 2024 20:28

Tgt Ion: 63 Resp: 78118
 Ion Ratio Lower Upper
 63 100
 98 6.5 3.3 9.8
 100 4.2 2.1 6.5









#25
2-Butanone
Concen: 260.137 ug/l
RT: 4.550 min Scan# 51
Delta R.T. -0.006 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 171196
Ion Ratio Lower Upper
43 100
72 25.8 19.4 29.2



#26
2,2-Dichloropropane
Concen: 46.682 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 77 Resp: 51827
Ion Ratio Lower Upper
77 100
97 23.9 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 53.847 ug/l

RT: 4.489 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 49216

Ion Ratio Lower Upper

96 100

61 144.3 0.0 312.2

98 63.5 0.0 129.0



#28

Bromochloromethane

Concen: 52.639 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28




Tgt Ion: 49 Resp: 32232


Ion Ratio Lower Upper

49 100

129 1.8 0.0 3.8

130 77.7 57.5 86.3





#29

Tetrahydrofuran

Concen: 271.444 ug/l

RT: 4.995 min Scan# 641

Delta R.T. -0.012 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 116381

Ion Ratio Lower Upper

42 100

72 44.8 34.2 51.4

71 42.0 32.2 48.2



#30

Chloroform

Concen: 54.786 ug/l

RT: 5.093 min Scan# 657

Delta R.T. -0.006 min

Lab File: VX040316.D




Acq: 17 Feb 2024 20:28


Tgt Ion: 83 Resp: 85428

Ion Ratio Lower Upper

83 100


85 64.8 50.3 75.5








#31
Cyclohexane
Concen: 50.007 ug/l
RT: 5.471 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 61143
Ion Ratio Lower Upper
56 100
69 33.1 23.4 35.2
84 87.0 65.9 98.9



#32
1,1,1-Trichloroethane
Concen: 53.409 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 97 Resp: 70062
Ion Ratio Lower Upper
97 100
99 65.6 51.1 76.7
61 46.8 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 50.141 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 51399

Ion Ratio Lower Upper

65 100

67 52.7 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.006 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28




Tgt Ion: 114 Resp: 116060


Ion Ratio Lower Upper

114 100

63 21.4 0.0 45.0

88 16.7 0.0 36.6









#35
Dibromofluoromethane
Concen: 51.244 ug/l
RT: 5.379 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 41533
Ion Ratio Lower Upper
113 100
111 102.8 82.9 124.3
192 18.6 13.3 19.9



#36
1,1-Dichloropropene
Concen: 51.456 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 75 Resp: 58422
Ion Ratio Lower Upper
75 100
110 36.3 17.0 50.9
77 30.7 24.7 37.1





#37

Ethyl Acetate

Concen: 53.705 ug/l

RT: 4.709 min Scan# 594

Delta R.T. -0.012 min


Lab File: VX040316.D

Acq: 17 Feb 2024 20:28




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:	43	Resp:	67847
Ion	Ratio	Lower	Upper
43	100		
61	13.8	10.6	16.0
70	9.9	8.3	12.5



#38

Carbon Tetrachloride

Concen: 52.766 ug/l




RT: 5.678 min Scan# 753


Delta R.T. 0.006 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Tgt Ion:	117	Resp:	56962
Ion	Ratio	Lower	Upper
117	100		
119	98.8	78.2	117.2
121	29.4	24.4	36.6









#39
Methylcyclohexane
Concen: 48.825 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 64423
Ion Ratio Lower Upper
83 100
55 72.3 64.6 97.0
98 45.6 36.7 55.1



#40
Benzene
Concen: 54.327 ug/l
RT: 6.038 min Scan# 812
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28


Tgt Ion: 78 Resp: 172490
Ion Ratio Lower Upper
78 100
77 25.4 19.3 28.9








#41
Methacrylonitrile
Concen: 53.625 ug/l
RT: 4.910 min Scan# 61
Delta R.T. -0.018 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41	Resp: 35914
Ion Ratio	Lower Upper
41 100	
39 56.3	45.2 67.8
67 78.2	52.6 78.8
52 34.4	22.7 34.1#



#42
1,2-Dichloroethane
Concen: 54.044 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 62	Resp: 67070
Ion Ratio	Lower Upper
62 100	
98 8.5	0.0 16.0





#43

Isopropyl Acetate

Concen: 54.283 ug/l

RT: 6.336 min Scan# 80

Delta R.T. -0.006 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 43 Resp: 107053

Ion Ratio Lower Upper

43 100

61 19.8 15.2 22.8

87 13.3 9.8 14.6



#44

Trichloroethene

Concen: 55.169 ug/l

RT: 7.123 min Scan# 990

Delta R.T. -0.006 min

Lab File: VX040316.D




Acq: 17 Feb 2024 20:28


Tgt Ion:130 Resp: 48317

Ion Ratio Lower Upper

130 100

95 103.2 0.0 204.4





#45

1,2-Dichloropropane

Concen: 56.494 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 46687

Ion Ratio Lower Upper

63 100

65 34.3 24.3 36.5



#46

Dibromomethane

Concen: 56.770 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28




Tgt Ion: 93 Resp: 34552


Ion Ratio Lower Upper

93 100

95 81.8 66.6 100.0

174 94.6 69.4 104.0









#47
Bromodichloromethane
Concen: 57.519 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.006 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 67454
Ion Ratio Lower Upper
83 100
85 62.9 49.8 74.6
127 8.2 7.1 10.7



#48
Methyl methacrylate
Concen: 53.968 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.006 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 41 Resp: 54459
Ion Ratio Lower Upper
41 100
69 84.0 59.2 88.8
39 62.7 47.7 71.5





#49

1,4-Dioxane

Concen: 1092.629 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 23986

Ion Ratio Lower Upper

88 100

43 32.1 0.0 0.0#

58 66.0 60.1 90.1



#50

Toluene-d8

Concen: 50.105 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.006 min

Lab File: VX040316.D




Acq: 17 Feb 2024 20:28


Tgt Ion: 98 Resp: 151910

Ion Ratio Lower Upper

98 100

100 64.9 53.1 79.7





#51

4-Methyl-2-Pentanone

Concen: 282.511 ug/l

RT: 8.574 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 380386

Ion Ratio Lower Upper

43 100

58 38.0 30.4 45.6



#52

Toluene

Concen: 54.146 ug/l

RT: 8.714 min Scan# 1251

Delta R.T. -0.006 min

Lab File: VX040316.D




Acq: 17 Feb 2024 20:28


Tgt Ion: 92 Resp: 112607

Ion Ratio Lower Upper

92 100

91 172.9 134.2 201.2









#53
t-1,3-Dichloropropene
Concen: 50.013 ug/l
RT: 8.976 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 69796
Ion Ratio Lower Upper
75 100
77 31.0 25.7 38.5



#54
cis-1,3-Dichloropropene
Concen: 55.717 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 75 Resp: 73229
Ion Ratio Lower Upper
75 100
77 32.2 25.3 37.9
39 48.4 46.7 70.1









#55
1,1,2-Trichloroethane
Concen: 55.251 ug/l
RT: 9.153 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 97	Resp: 49008
Ion Ratio	Lower Upper
97 100	
83 88.4	65.8 98.8
85 59.4	41.6 62.4
99 62.8	47.8 71.8



#56
Ethyl methacrylate
Concen: 55.889 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 69	Resp: 74215
Ion Ratio	Lower Upper
69 100	
41 66.8	56.8 85.2
39 43.2	32.0 48.0









#57
1,3-Dichloropropane
Concen: 55.669 ug/l
RT: 9.305 min Scan# 1134
Delta R.T. -0.006 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 80598
Ion Ratio Lower Upper
76 100
78 32.2 26.1 39.1



#58
2-Chloroethyl Vinyl ether
Concen: 255.169 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. -0.006 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 63 Resp: 160290
Ion Ratio Lower Upper
63 100
106 27.2 22.5 33.7









#59
2-Hexanone
Concen: 276.977 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 295681
Ion Ratio Lower Upper
43 100
58 53.5 26.2 78.5



#60
Dibromochloromethane
Concen: 55.946 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 129 Resp: 50559
Ion Ratio Lower Upper
129 100
127 78.1 38.8 116.4









#61
1,2-Dibromoethane
Concen: 55.035 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 51889
Ion Ratio Lower Upper
107 100
109 94.7 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 51.287 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 95 Resp: 63417
Ion Ratio Lower Upper
95 100
174 76.6 0.0 135.6
176 72.3 0.0 131.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:117 Resp: 113530

Ion Ratio Lower Upper

117 100

82 55.0 46.2 69.4

119 31.8 25.5 38.3



#64

Tetrachloroethene

Concen: 51.785 ug/l

RT: 9.275 min Scan# 1343

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Tgt Ion:164 Resp: 41199




Ion Ratio Lower Upper


164 100

166 124.1 102.7 154.1

129 98.5 80.4 120.6

131 92.4 76.5 114.7





#65

Chlorobenzene

Concen: 52.240 ug/l

RT: 10.080 min Scan# 1475

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:112 Resp: 123778

Ion Ratio Lower Upper

112 100

114 31.7 25.7 38.5



#66

1,1,1,2-Tetrachloroethane

Concen: 55.839 ug/l

RT: 10.165 min Scan# 1489

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28




Tgt Ion:131 Resp: 45823


Ion Ratio Lower Upper

131 100

133 93.3 47.1 141.4

119 64.7 32.2 96.6





#67

Ethyl Benzene

Concen: 53.766 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.006 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 226888

Ion Ratio Lower Upper

91 100

106 30.1 24.9 37.3



#68

m/p-Xylenes

Concen: 106.957 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX040316.D




Acq: 17 Feb 2024 20:28


Tgt Ion:106 Resp: 170463

Ion Ratio Lower Upper

106 100

91 208.9 165.5 248.3









#69
o-Xylene
Concen: 54.147 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 83861
Ion Ratio Lower Upper
106 100
91 221.1 108.9 326.8



#70
Styrene
Concen: 55.886 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion:104 Resp: 143903
Ion Ratio Lower Upper
104 100
78 52.3 43.6 65.4
103 55.5 43.6 65.4





#71

Bromoform

Concen: 49.396 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28





Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:173 Resp: 38053

Ion	Ratio	Lower	Upper
173	100		
175	49.0	24.1	72.4
254	0.1	0.0	0.0#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794




Delta R.T. 0.000 min


Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Tgt Ion:152 Resp: 61912

Ion	Ratio	Lower	Upper
152	100		
115	86.4	43.3	129.9
150	175.1	0.0	345.0









#73
Isopropylbenzene
Concen: 52.799 ug/l
RT: 10.964 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 230237
Ion Ratio Lower Upper
105 100
120 25.9 12.9 38.6



#74
N-ethyl acetate
Concen: 55.471 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 43 Resp: 103683
Ion Ratio Lower Upper
43 100
70 40.4 32.9 49.3
55 24.1 19.3 28.9
61 23.1 18.8 28.2









#75
1,1,2,2-Tetrachloroethane
Concen: 54.603 ug/l
RT: 11.214 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 85545
Ion Ratio Lower Upper
83 100
131 9.5 5.3 15.8
85 63.2 32.0 96.2



#76
1,2,3-Trichloropropane
Concen: 71.401 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX040316.D
Acq: 17 Feb 2024 20:28

Tgt Ion: 75 Resp: 89519
Ion Ratio Lower Upper
75 100
77 31.8 19.1 57.5





#77

Bromobenzene

Concen: 53.350 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.006 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:156 Resp: 54947

Ion Ratio Lower Upper

156 100

77 169.6 88.4 265.3

158 96.2 48.0 144.0



#78

n-propylbenzene

Concen: 52.714 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.000 min

Lab File: VX040316.D




Acq: 17 Feb 2024 20:28


Tgt Ion: 91 Resp: 278297

Ion Ratio Lower Upper

91 100

120 22.1 11.5 34.4





#79

2-Chlorotoluene

Concen: 51.061 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 162779

Ion Ratio Lower Upper

91 100

126 33.8 16.3 48.8



#80

1,3,5-Trimethylbenzene

Concen: 52.977 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX040316.D




Acq: 17 Feb 2024 20:28


Tgt Ion: 105 Resp: 194488

Ion Ratio Lower Upper

105 100

120 46.3 24.1 72.3









#81
 trans-1,4-Dichloro-2-butene
 Concen: 46.953 ug/l
 RT: 11.018 min Scan# 1
 Delta R.T. 0.000 min
 Lab File: VX040316.D
 Acq: 17 Feb 2024 20:28




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 20859
 Ion Ratio Lower Upper
 75 100
 53 106.9 90.5 135.7
 89 45.8 36.9 55.3



#82
 4-Chlorotoluene
 Concen: 51.195 ug/l
 RT: 11.451 min Scan# 1700
 Delta R.T. 0.000 min
 Lab File: VX040316.D
 Acq: 17 Feb 2024 20:28

Tgt Ion: 91 Resp: 193374
 Ion Ratio Lower Upper
 91 100
 126 28.8 14.2 42.6





#83

tert-Butylbenzene

Concen: 51.676 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:119 Resp: 185994

Ion Ratio Lower Upper

119 100

91 63.3 31.2 93.6

134 22.8 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 53.042 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX040316.D




Acq: 17 Feb 2024 20:28


Tgt Ion:105 Resp: 197006

Ion Ratio Lower Upper

105 100

120 42.7 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 53.262 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX040316.D
 Acq: 17 Feb 2024 20:28




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 246364
 Ion Ratio Lower Upper
 105 100
 134 19.7 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 52.938 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX040316.D
 Acq: 17 Feb 2024 20:28

Tgt Ion:119 Resp: 200722
 Ion Ratio Lower Upper
 119 100
 134 26.7 12.8 38.3
 91 25.8 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 52.347 ug/l

RT: 11.970 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 106824

Ion Ratio Lower Upper

146 100

111 41.9 22.1 66.1

148 63.8 31.8 95.3



#88

1,4-Dichlorobenzene

Concen: 50.253 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28




Tgt Ion:146 Resp: 108592


Ion Ratio Lower Upper

146 100

111 40.1 21.3 63.7

148 63.9 31.8 95.4





#89

n-Butylbenzene

Concen: 51.931 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.006 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 190072

Ion Ratio Lower Upper

91 100

92 52.9 27.4 82.0

134 24.5 12.6 37.8



#90

Hexachloroethane

Concen: 54.343 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX040316.D




Acq: 17 Feb 2024 20:28


Tgt Ion: 117 Resp: 32501

Ion Ratio Lower Upper

117 100

201 72.0 32.9 98.6









#91
 1,2-Dichlorobenzene
 Concen: 53.592 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.000 min
 Lab File: VX040316.D
 Acq: 17 Feb 2024 20:28




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:146 Resp: 105871
 Ion Ratio Lower Upper
 146 100
 111 43.0 22.4 67.0
 148 63.4 31.6 95.0



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 54.861 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. -0.006 min
 Lab File: VX040316.D
 Acq: 17 Feb 2024 20:28

Tgt Ion: 75 Resp: 19837
 Ion Ratio Lower Upper
 75 100
 155 75.9 36.1 108.5
 157 97.2 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 50.309 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 68042

Ion Ratio Lower Upper

180 100

182 95.4 48.3 144.8

145 31.5 18.0 54.0



#94

Hexachlorobutadiene

Concen: 49.059 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28




Tgt Ion:225 Resp: 26481


Ion Ratio Lower Upper

225 100

223 64.4 31.1 93.2

227 62.5 32.5 97.5





#95

Naphthalene

Concen: 52.664 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28





Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:128 Resp: 224958

Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.3	15.5
129	10.9	8.8	13.2



#96

1,2,3-Trichlorobenzene

Concen: 52.840 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.000 min

Lab File: VX040316.D

Acq: 17 Feb 2024 20:28

Tgt Ion:180 Resp: 68120

Ion	Ratio	Lower	Upper
180	100		
182	94.0	46.9	140.6
145	32.1	17.8	53.5

