

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
 Data File : VX040300.D
 Acq On : 17 Feb 2024 14:24
 Operator : JC/MD
 Sample : P1507-02DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D4-20240212DL

Quant Time: Feb 19 03:57:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	77442	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	131033	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64572	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	52629	45.681	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.360%
35) Dibromofluoromethane	5.385	113	42685	46.647	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.300%
50) Toluene-d8	8.647	98	157428	45.992	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.980%#
62) 4-Bromofluorobenzene	11.079	95	62179	44.540	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.080%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	29187	35.543	ug/l	96
44) Trichloroethene	7.129	130	4534	4.585	ug/l	80
64) Tetrachloroethene	9.268	164	3092	3.467	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
Data File : VX040300.D
Acq On : 17 Feb 2024 14:24
Operator : JC/MD
Sample : P1507-02DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D4-20240212DL

Quant Time: Feb 19 03:57:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

