

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
 Data File : VX040301.D
 Acq On : 17 Feb 2024 14:47
 Operator : JC/MD
 Sample : P1507-13DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D3-20240214DL

Quant Time: Feb 19 03:57:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77470	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	134389	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66454	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	53520	46.437	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.880%
35) Dibromofluoromethane	5.391	113	43920	46.798	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.600%
50) Toluene-d8	8.647	98	162256	46.218	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.440%
62) 4-Bromofluorobenzene	11.079	95	63882	44.617	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.240%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	36210	44.079	ug/l	97
44) Trichloroethene	7.129	130	4617	4.553	ug/l	100
64) Tetrachloroethene	9.275	164	2943	3.266	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
Data File : VX040301.D
Acq On : 17 Feb 2024 14:47
Operator : JC/MD
Sample : P1507-13DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D3-20240214DL

Quant Time: Feb 19 03:57:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

