

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
 Data File : VX040302.D
 Acq On : 17 Feb 2024 15:09
 Operator : JC/MD
 Sample : P1507-19DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D2-20240215DL

Quant Time: Feb 19 03:57:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	71577	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	122474	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118319	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	49136	46.143	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.280%
35) Dibromofluoromethane	5.385	113	39452	46.127	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.260%
50) Toluene-d8	8.647	98	147902	46.228	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.460%
62) 4-Bromofluorobenzene	11.079	95	57257	43.881	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.760%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	28949	31.323	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
Data File : VX040302.D
Acq On : 17 Feb 2024 15:09
Operator : JC/MD
Sample : P1507-19DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D2-20240215DL

Quant Time: Feb 19 03:57:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

