

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
 Data File : VX040303.D
 Acq On : 17 Feb 2024 15:32
 Operator : JC/MD
 Sample : P1507-28DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20240212DL

Quant Time: Feb 19 03:58:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	70478	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	121484	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120775	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62945	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	48590	46.342	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.680%
35) Dibromofluoromethane	5.385	113	39501	46.561	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.120%
50) Toluene-d8	8.647	98	148905	46.921	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.840%
62) 4-Bromofluorobenzene	11.079	95	58590	45.268	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.540%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	31519	42.175	ug/l	Qvalue 96
44) Trichloroethene	7.129	130	7911	8.630	ug/l	85
64) Tetrachloroethene	9.275	164	10408	12.297	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
Data File : VX040303.D
Acq On : 17 Feb 2024 15:32
Operator : JC/MD
Sample : P1507-28DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240212DL

Quant Time: Feb 19 03:58:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

