

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
 Data File : VX040305.D
 Acq On : 17 Feb 2024 16:18
 Operator : JC/MD
 Sample : P1519-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Feb 19 03:58:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	72126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	120289	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119316	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60865	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	48405	45.111	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.220%
35) Dibromofluoromethane	5.385	113	39616	47.160	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.320%
50) Toluene-d8	8.647	98	148331	47.204	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.400%
62) 4-Bromofluorobenzene	11.079	95	57173	44.612	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
Data File : VX040305.D
Acq On : 17 Feb 2024 16:18
Operator : JC/MD
Sample : P1519-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Feb 19 03:58:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

