

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
 Data File : VX040311.D
 Acq On : 17 Feb 2024 18:35
 Operator : JC/MD
 Sample : P1506-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-840-842

Quant Time: Feb 19 04:00:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70423	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	120827	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59951	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	46905	44.770	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.540%
35) Dibromofluoromethane	5.385	113	38995	46.214	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.420%
50) Toluene-d8	8.647	98	144008	45.625	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.240%#
62) 4-Bromofluorobenzene	11.079	95	56347	43.772	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.540%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1178	2.738	ug/l	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
Data File : VX040311.D
Acq On : 17 Feb 2024 18:35
Operator : JC/MD
Sample : P1506-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-840-842

Quant Time: Feb 19 04:00:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

