

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
 Data File : VX040313.D
 Acq On : 17 Feb 2024 19:20
 Operator : JC/MD
 Sample : P1507-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D2-20240215

Quant Time: Feb 19 04:01:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70148	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	120415	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	47645	45.655	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.300%
35) Dibromofluoromethane	5.385	113	39607	47.100	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.200%
50) Toluene-d8	8.647	98	143786	45.710	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.420%#
62) 4-Bromofluorobenzene	11.079	95	55522	43.278	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.560%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	1662	2.321	ug/l	82
27) cis-1,2-Dichloroethene	4.489	96	2340	2.515	ug/l	83
32) 1,1,1-Trichloroethane	5.379	97	775	0.580	ug/l	# 49
44) Trichloroethene	7.129	130	29322	32.269	ug/l	94
64) Tetrachloroethene	9.275	164	2859	3.503	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021724\
Data File : VX040313.D
Acq On : 17 Feb 2024 19:20
Operator : JC/MD
Sample : P1507-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D2-20240215

Quant Time: Feb 19 04:01:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

