

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021819\
 Data File : VX007578.D
 Acq On : 18 Feb 2019 10:30
 Operator : JC/SP
 Sample : VX0218WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0218WBL01

Quant Time: Feb 19 04:36:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	533489	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	819604	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	752385	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	358248	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	282773	49.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.88%	
35) Dibromofluoromethane	5.50	113	268429	50.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.66%	
50) Toluene-d8	8.71	98	974424	49.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.50%	
62) 4-Bromofluorobenzene	11.13	95	330421	45.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.48%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021819\
Data File : VX007578.D
Acq On : 18 Feb 2019 10:30
Operator : JC/SP
Sample : VX0218WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0218WBL01

Quant Time: Feb 19 04:36:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

