

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021819\
 Data File : VX007587.D
 Acq On : 18 Feb 2019 14:38
 Operator : JC/SP
 Sample : PB117159TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB117159TB

Quant Time: Feb 19 04:59:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	469432	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	749472	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	721605	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	338823	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	264095	53.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.00%	
35) Dibromofluoromethane	5.51	113	246841	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
50) Toluene-d8	8.71	98	919504	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
62) 4-Bromofluorobenzene	11.13	95	315263	47.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.46%	
Target Compounds						
16) Acetone	2.45	43	31066	13.611	ug/l	100
43) Isopropyl Acetate	6.46	43	25031	2.386	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021819\
Data File : VX007587.D
Acq On : 18 Feb 2019 14:38
Operator : JC/SP
Sample : PB117159TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB117159TB

Quant Time: Feb 19 04:59:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

