

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021819\
 Data File : VX007588.D
 Acq On : 18 Feb 2019 15:05
 Operator : JC/SP
 Sample : PB117159ZHE#19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB117159ZHE#19

Quant Time: Feb 19 05:04:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	472856	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	761135	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	728897	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	340040	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	270299	53.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.72%	
35) Dibromofluoromethane	5.51	113	251835	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
50) Toluene-d8	8.71	98	929625	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
62) 4-Bromofluorobenzene	11.13	95	318968	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
Target Compounds						
16) Acetone	2.45	43	30885	13.434	ug/l	100
18) Methyl Acetate	2.78	43	18257	3.605	ug/l	93
20) Methylene Chloride	2.86	84	8026	1.326	ug/l	94
43) Isopropyl Acetate	6.46	43	24750	2.323	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021819\
Data File : VX007588.D
Acq On : 18 Feb 2019 15:05
Operator : JC/SP
Sample : PB117159ZHE#19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB117159ZHE#19

Quant Time: Feb 19 05:04:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

