

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021819\
 Data File : VX007590.D
 Acq On : 18 Feb 2019 15:59
 Operator : JC/SP
 Sample : K1508-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M-12-S

Quant Time: Feb 19 05:10:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	457725	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	744870	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	717254	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	341035	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	262703	54.07	ug/l	0.00
Spiked Amount			Recovery	=	108.14%	
35) Dibromofluoromethane	5.50	113	245610	50.68	ug/l	0.00
Spiked Amount			Recovery	=	101.36%	
50) Toluene-d8	8.71	98	913516	50.80	ug/l	0.00
Spiked Amount			Recovery	=	101.60%	
62) 4-Bromofluorobenzene	11.13	95	318730	48.55	ug/l	0.00
Spiked Amount			Recovery	=	97.10%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	16579	7.450	ug/l	95
18) Methyl Acetate	2.78	43	17602	3.593	ug/l	98
20) Methylene Chloride	2.86	84	340896	75.190	ug/l	98
43) Isopropyl Acetate	6.46	43	25708	2.465	ug/l	97
52) Toluene	8.79	92	50337	3.963	ug/l	98
67) Ethyl Benzene	10.25	91	53300	2.115	ug/l	99
68) m/p-Xylenes	10.36	106	85268	8.510	ug/l	100
69) o-Xylene	10.70	106	39092	4.055	ug/l	99
78) n-propylbenzene	11.36	91	27155	1.069	ug/l	97
80) 1,3,5-Trimethylbenzene	11.51	105	53058	2.798	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	187359	9.733	ug/l	99
95) Naphthalene	13.83	128	84326	3.666	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021819\
Data File : VX007590.D
Acq On : 18 Feb 2019 15:59
Operator : JC/SP
Sample : K1508-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M-12-S

Quant Time: Feb 19 05:10:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

