

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021819\
 Data File : VX007594.D
 Acq On : 18 Feb 2019 17:45
 Operator : JC/SP
 Sample : K1511-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STONES

Quant Time: Feb 19 05:52:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	445009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	713044	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	699700	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	328679	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	249692	52.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.72%	
35) Dibromofluoromethane	5.51	113	234773	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
50) Toluene-d8	8.71	98	883390	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
62) 4-Bromofluorobenzene	11.13	95	311249	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
Target Compounds						
16) Acetone	2.45	43	16676	7.707	ug/l	94
18) Methyl Acetate	2.78	43	18198	3.783	ug/l	94
20) Methylene Chloride	2.86	84	90363	20.212	ug/l	97
43) Isopropyl Acetate	6.46	43	24149	2.419	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021819\
Data File : VX007594.D
Acq On : 18 Feb 2019 17:45
Operator : JC/SP
Sample : K1511-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STONES

Quant Time: Feb 19 05:52:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

