

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021819\
 Data File : VX007597.D
 Acq On : 18 Feb 2019 19:05
 Operator : JC/SP
 Sample : K1513-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 INNER-WALL

Quant Time: Feb 19 05:59:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	449987	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	719469	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	681710	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	318979	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	251582	52.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.34%	
35) Dibromofluoromethane	5.51	113	235606	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
50) Toluene-d8	8.71	98	873659	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
62) 4-Bromofluorobenzene	11.13	95	304049	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
Target Compounds						
16) Acetone	2.45	43	19491	8.909	ug/l	97
18) Methyl Acetate	2.78	43	6983	1.804	ug/l	94
20) Methylene Chloride	2.86	84	5634199	1270.357	ug/l	98
25) 2-Butanone	4.70	43	9801	3.148	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021819\
Data File : VX007597.D
Acq On : 18 Feb 2019 19:05
Operator : JC/SP
Sample : K1513-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
INNER-WALL

Quant Time: Feb 19 05:59:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

