

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021819\
 Data File : VX007598.D
 Acq On : 18 Feb 2019 19:32
 Operator : JC/SP
 Sample : K1517-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 G-021519

Manual Integrations
 APPROVED

MMDadoda
 2/19/2019 11:01:23 AM

Quant Time: Feb 19 06:02:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	444073	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	719353	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	696289	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	330086	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	253512	53.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.56%	
35) Dibromofluoromethane	5.50	113	234987	50.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.40%	
50) Toluene-d8	8.71	98	886204	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
62) 4-Bromofluorobenzene	11.13	95	309316	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
Target Compounds						
16) Acetone	2.45	43	17461	8.087	ug/l	95
20) Methylene Chloride	2.86	84	22449	4.734	ug/l	96
25) 2-Butanone	4.70	43	6593	2.146	ug/l	98
43) Isopropyl Acetate	6.46	43	22184	2.203	ug/l	96
95) Naphthalene	13.83	128	170150m	7.643	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021819\
Data File : VX007598.D
Acq On : 18 Feb 2019 19:32
Operator : JC/SP
Sample : K1517-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
G-021519

Manual Integrations
APPROVED

MMDadoda
2/19/2019 11:01:23 AM

Quant Time: Feb 19 06:02:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

