

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021819\
 Data File : VX007599.D
 Acq On : 18 Feb 2019 19:58
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 19 06:03:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.557	0.516	7.4	72	0.00
3 P	Chloromethane	0.563	0.491	12.8	78	0.00
4 C	Vinyl Chloride	0.552	0.524	5.1#	80	0.00
5 T	Bromomethane	0.629	0.496	21.1#	77	0.00
6 T	Chloroethane	0.357	0.341	4.5	87	0.00
7 T	Trichlorofluoromethane	0.864	0.825	4.5	85	0.00
8 T	Diethyl Ether	0.325	0.324	0.3	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.514	0.453	11.9	81	0.00
10 T	Methyl Iodide	0.756	0.663	12.3	74	0.00
11 T	Tert butyl alcohol	0.110	0.113	-2.7	98	0.00
12 CM	1,1-Dichloroethene	0.477	0.427	10.5#	83	0.00
13 T	Acrolein	0.052	0.038	26.9#	64	0.00
14 T	Allyl chloride	0.778	0.729	6.3	84	0.00
15 T	Acrylonitrile	0.272	0.272	0.0	96	0.00
16 T	Acetone	0.243	0.235	3.3	91	0.00
17 T	Carbon Disulfide	1.260	0.985	21.8#	73	0.00
18 T	Methyl Acetate	0.688	0.764	-11.0	109	0.00
19 T	Methyl tert-butyl Ether	1.547	1.515	2.1	92	0.00
20 T	Methylene Chloride	0.550	0.497	9.6	89	0.00
21 T	trans-1,2-Dichloroethene	0.509	0.450	11.6	83	0.00
22 T	Diisopropyl ether	1.431	1.466	-2.4	93	0.00
23 T	Vinyl Acetate	1.187	1.214	-2.3	91	0.00
24 P	1,1-Dichloroethane	0.822	0.822	0.0	91	0.00
25 T	2-Butanone	0.346	0.379	-9.5	103	0.01
26 T	2,2-Dichloropropane	0.626	0.567	9.4	80	0.00
27 T	cis-1,2-Dichloroethene	0.555	0.553	0.4	94	0.00
28 T	Bromochloromethane	0.378	0.398	-5.3	92	0.00
29 T	Tetrahydrofuran	0.218	0.247	-13.3	105	0.00
30 C	Chloroform	0.857	0.901	-5.1#	96	0.00
31 T	Cyclohexane	0.738	0.679	8.0	83	0.00
32 T	1,1,1-Trichloroethane	0.724	0.765	-5.7	93	0.00
33 S	1,2-Dichloroethane-d4	0.531	0.550	-3.6	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.325	0.340	-4.6	96	0.00
36 T	1,1-Dichloropropene	0.449	0.406	9.6	86	0.00
37 T	Ethyl Acetate	0.453	0.474	-4.6	102	0.00
38 T	Carbon Tetrachloride	0.435	0.444	-2.1	91	0.00
39 T	Methylcyclohexane	0.543	0.464	14.5	79	0.00
40 TM	Benzene	1.321	1.291	2.3	92	0.00
41 T	Methacrylonitrile	0.243	0.263	-8.2	100	0.00
42 TM	1,2-Dichloroethane	0.457	0.447	2.2	93	0.00
43 T	Isopropyl Acetate	0.700	0.758	-8.3	100	0.01
44 TM	Trichloroethene	0.398	0.373	6.3	88	0.00
45 C	1,2-Dichloropropane	0.331	0.333	-0.6#	94	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021819\
 Data File : VX007599.D
 Acq On : 18 Feb 2019 19:58
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 19 06:03:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.241	0.239	0.8	96	0.00
47 T	Bromodichloromethane	0.408	0.441	-8.1	99	0.00
48 T	Methyl methacrylate	0.350	0.385	-10.0	98	0.00
49 T	1,4-Dioxane	0.009	0.008	11.1	99	0.00
50 S	Toluene-d8	1.207	1.218	-0.9	91	0.00
51 T	4-Methyl-2-Pentanone	0.461	0.505	-9.5	100	0.00
52 CM	Toluene	0.853	0.827	3.0#	90	0.00
53 T	t-1,3-Dichloropropene	0.438	0.464	-5.9	91	0.00
54 T	cis-1,3-Dichloropropene	0.482	0.511	-6.0	92	0.00
55 T	1,1,2-Trichloroethane	0.355	0.366	-3.1	98	0.00
56 T	Ethyl methacrylate	0.492	0.523	-6.3	95	0.00
57 T	1,3-Dichloropropane	0.565	0.573	-1.4	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.279	-7.7	95	0.00
59 T	2-Hexanone	0.359	0.381	-6.1	99	0.00
60 T	Dibromochloromethane	0.336	0.389	-15.8	97	0.00
61 T	1,2-Dibromoethane	0.376	0.390	-3.7	95	0.00
62 S	4-Bromofluorobenzene	0.441	0.439	0.5	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.471	0.417	11.5	84	0.00
65 PM	Chlorobenzene	1.091	1.045	4.2	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.366	0.396	-8.2	96	0.00
67 C	Ethyl Benzene	1.757	1.728	1.7#	90	0.00
68 T	m/p-Xylenes	0.698	0.673	3.6	87	0.00
69 T	o-Xylene	0.672	0.655	2.5	88	0.00
70 T	Styrene	1.090	1.083	0.6	89	0.00
71 P	Bromoform	0.283	0.317	-12.0	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	3.398	3.435	-1.1	89	0.00
74 T	N-amyl acetate	1.314	1.408	-7.2	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.036	1.120	-8.1	99	0.00
76 T	1,2,3-Trichloropropane	0.973	1.065	-9.5	95	0.00
77 T	Bromobenzene	0.957	0.928	3.0	87	0.00
78 T	n-propylbenzene	3.726	3.798	-1.9	87	0.00
79 T	2-Chlorotoluene	2.261	2.232	1.3	87	0.00
80 T	1,3,5-Trimethylbenzene	2.780	2.828	-1.7	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.280	0.315	-12.5	93	0.00
82 T	4-Chlorotoluene	2.616	2.614	0.1	86	0.00
83 T	tert-Butylbenzene	2.752	2.790	-1.4	89	0.00
84 T	1,2,4-Trimethylbenzene	2.822	2.880	-2.1	87	0.00
85 T	sec-Butylbenzene	3.331	3.433	-3.1	87	0.00
86 T	p-Isopropyltoluene	3.007	3.064	-1.9	86	0.00
87 T	1,3-Dichlorobenzene	1.718	1.635	4.8	85	0.00
88 T	1,4-Dichlorobenzene	1.743	1.634	6.3	85	0.00
89 T	n-Butylbenzene	2.587	2.549	1.5	83	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021819\
Data File : VX007599.D
Acq On : 18 Feb 2019 19:58
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 19 06:03:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.450	0.516	-14.7	95	0.00
91 T	1,2-Dichlorobenzene	1.710	1.656	3.2	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.215	0.242	-12.6	101	0.00
93 T	1,2,4-Trichlorobenzene	1.153	1.113	3.5	85	0.00
94 T	Hexachlorobutadiene	0.560	0.545	2.7	88	0.00
95 T	Naphthalene	3.372	3.535	-4.8	92	0.00
96 T	1,2,3-Trichlorobenzene	1.161	1.134	2.3	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6