

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021819\
 Data File : VX007583.D
 Acq On : 18 Feb 2019 12:52
 Operator : JC/SP
 Sample : K1507-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Feb 19 04:50:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	476623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	755876	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	696457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	325971	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	264071	52.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.40%	
35) Dibromofluoromethane	5.50	113	247384	50.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.60%	
50) Toluene-d8	8.71	98	898642	49.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.50%	
62) 4-Bromofluorobenzene	11.13	95	311395	46.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.48%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021819\
Data File : VX007583.D
Acq On : 18 Feb 2019 12:52
Operator : JC/SP
Sample : K1507-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Feb 19 04:50:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.01 min

Lab File: VX007583.D

Acq: 18 Feb 2019 12:52

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-SAMPLE-MANA

Tot Ion:168 Resp: 476623

Ion Ratio Lower Upper

168 100

99 51.0 37.9 56.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67



150000

100000

50000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 52.199 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX007583.D


Acq: 18 Feb 2019 12:52

Tgt Ion: 65 Resp: 264071

Ion Ratio Lower Upper

65 100

67 53.5 0.0 108.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

120000

100000

80000


60000


40000

20000

0

Time--> 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX007583.D

Acq: 18 Feb 2019 12:52

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-SAMPLE-MANA


Tot Ion:114 Resp: 755876

Ion Ratio Lower Upper

114 100

63 18.1 0.0 35.6




88 15.3 0.0 29.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 50.300 ug/l

RT: 5.50 min Scan# 713

Delta R.T. -0.01 min

Lab File: VX007583.D

Acq: 18 Feb 2019 12:52


Tgt Ion:113 Resp: 247384

Ion Ratio Lower Upper

113 100

111 102.5 81.8 122.8



192 22.3 18.5 27.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.248 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX007583.D

Acq: 18 Feb 2019 12:52

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-SAMPLE-MANA

Tot Ion: 98 Resp: 898642

Ion Ratio Lower Upper




98 100

100 64.4 51.6 77.4



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 46.741 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX007583.D

Acq: 18 Feb 2019 12:52


Tgt Ion: 95 Resp: 311395

Ion Ratio Lower Upper

95 100

174 93.6 0.0 189.2



176 89.4 0.0 186.2




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. 0.01 min
Lab File: VX007583.D
Acq: 18 Feb 2019 12:52

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Tot Ion	Ratio	Lower	Upper
117	100		
82	50.1	42.1	63.1
119	32.7	26.5	39.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX007583.D
Acq: 18 Feb 2019 12:52

Tgt Ion	Ratio	Lower	Upper
152	100		
115	55.8	38.0	114.0
150	159.6	0.0	346.4

