


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021819\
Data File : VX007576.D
Acq On : 18 Feb 2019 08:30
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Title : SW846 8260
Last Update : Sat Feb 02 01:32:36 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1624

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	8235	PASS
75	95	30	60	49.9	21780	PASS
95	95	100	100	100.0	43656	PASS
96	95	5	9	6.8	2986	PASS
173	174	0.00	2	1.0	397	PASS
174	95	50	100	92.9	40576	PASS
175	174	5	9	7.8	3170	PASS
176	174	95	101	97.0	39354	PASS
177	176	5	9	6.3	2461	PASS