

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021820\
 Data File : VX014952.D
 Acq On : 18 Feb 2020 12:23
 Operator : JC/SP
 Sample : L1564-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS027-200217

Quant Time: Feb 19 00:48:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	433765	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	678518	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	641303	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	332929	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	244622	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
35) Dibromofluoromethane	5.49	113	206695	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
50) Toluene-d8	8.71	98	805841	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
62) 4-Bromofluorobenzene	11.14	95	297343	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	218090	44.472	ug/l	89
44) Trichloroethene	7.21	130	20097	3.817	ug/l	99
64) Tetrachloroethene	9.34	164	13768	2.296	ug/l	94
65) Chlorobenzene	10.13	112	19411	1.492	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021820\
Data File : VX014952.D
Acq On : 18 Feb 2020 12:23
Operator : JC/SP
Sample : L1564-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS027-200217

Quant Time: Feb 19 00:48:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

