

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021820\
 Data File : VX014953.D
 Acq On : 18 Feb 2020 12:46
 Operator : JC/SP
 Sample : L1564-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS037-200217

Quant Time: Feb 19 00:49:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	426071	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	667331	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	640971	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	339735	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	242772	50.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.98%	
35) Dibromofluoromethane	5.49	113	204562	49.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.72%	
50) Toluene-d8	8.71	98	800221	51.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.20%	
62) 4-Bromofluorobenzene	11.14	95	301632	53.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.94%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021820\
Data File : VX014953.D
Acq On : 18 Feb 2020 12:46
Operator : JC/SP
Sample : L1564-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS037-200217

Quant Time: Feb 19 00:49:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

