

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
 Data File : VX027097.D
 Acq On : 17 Feb 2022 23:54
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 18 05:14:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X021822W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Feb 18 05:14:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.910	128	14439	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	79189	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	68101	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	31180	29.865	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.533%
60) 4-Bromofluorobenzene	11.085	95	30241	27.308	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.033%
63) Toluene-d8	8.653	98	91742	30.047	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.167%
Target Compounds						
16) Carbon Disulfide	2.514	76	3820	1.741	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
Data File : VX027097.D
Acq On : 17 Feb 2022 23:54
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 18 05:14:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X021822W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Feb 18 05:14:00 2022
Response via : Initial Calibration

