

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
 Data File : VX027107.D
 Acq On : 18 Feb 2022 14:07
 Operator : JC/MD
 Sample : N1583-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 220216008-06-TRIP-BLANK

Quant Time: Feb 21 00:21:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X021822W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Feb 18 05:14:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.910	128	14445	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	80004	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	68595	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	31137	29.811	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.367%
60) 4-Bromofluorobenzene	11.079	95	31390	28.141	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.800%
63) Toluene-d8	8.653	98	92615	30.115	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.367%
Target Compounds						
15) Acetone	2.379	58	479	3.299	ug/l	Qvalue # 49

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
Data File : VX027107.D
Acq On : 18 Feb 2022 14:07
Operator : JC/MD
Sample : N1583-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
220216008-06-TRIP-BLANK

Quant Time: Feb 21 00:21:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X021822W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Feb 18 05:14:00 2022
Response via : Initial Calibration

