

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
 Data File : VX027109.D
 Acq On : 18 Feb 2022 14:53
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 21 00:38:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	114774	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	190712	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	165259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70908	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69281	50.286	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.580%	
35) Dibromofluoromethane	5.391	113	58626	51.597	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.200%	
50) Toluene-d8	8.653	98	215512	51.672	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.340%	
62) 4-Bromofluorobenzene	11.079	95	75569	51.464	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.920%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
Data File : VX027109.D
Acq On : 18 Feb 2022 14:53
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 21 00:38:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

