

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
 Data File : VX027112.D
 Acq On : 18 Feb 2022 16:06
 Operator : JC/MD
 Sample : N1539-03ME
 Misc : 6.14g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PE-3ME

Quant Time: Feb 21 00:39:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	96920	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	158957	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	140312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	57478	49.404	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.800%
35) Dibromofluoromethane	5.397	113	47060	49.692	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.380%
50) Toluene-d8	8.653	98	180488	51.920	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.840%
62) 4-Bromofluorobenzene	11.085	95	67169	54.882	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.760%
Target Compounds						
					Qvalue	
52) Toluene	8.720	92	10559	3.830	ug/l	98
67) Ethyl Benzene	10.195	91	12091	2.354	ug/l	94
68) m/p-Xylenes	10.305	106	14949	7.565	ug/l	99
69) o-Xylene	10.647	106	9844	5.084	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	16222	3.808	ug/l	98
95) Naphthalene	13.780	128	7983	1.674	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
Data File : VX027112.D
Acq On : 18 Feb 2022 16:06
Operator : JC/MD
Sample : N1539-03ME
Misc : 6.14g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PE-3ME

Quant Time: Feb 21 00:39:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

