

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
 Data File : VX027113.D
 Acq On : 18 Feb 2022 16:29
 Operator : JC/MD
 Sample : N1539-01 20X
 Misc : 4.00g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PE-1

Quant Time: Feb 21 00:40:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	102652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	170447	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150248	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67427	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62480	50.705	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.400%
35) Dibromofluoromethane	5.391	113	52512	51.711	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.420%
50) Toluene-d8	8.653	98	193895	52.016	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.040%
62) 4-Bromofluorobenzene	11.079	95	71570	54.536	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
Data File : VX027113.D
Acq On : 18 Feb 2022 16:29
Operator : JC/MD
Sample : N1539-01 20X
Misc : 4.00g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PE-1

Quant Time: Feb 21 00:40:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

