

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
 Data File : VX027116.D
 Acq On : 18 Feb 2022 17:39
 Operator : JC/MD
 Sample : N1465-02ME
 Misc : 4.90g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A5-2-6.5ME

Quant Time: Feb 21 04:04:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	100743	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	165772	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	151143	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74323	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59510	49.210	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.420%
35) Dibromofluoromethane	5.391	113	48243	48.847	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.700%
50) Toluene-d8	8.653	98	190742	52.614	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.220%
62) 4-Bromofluorobenzene	11.085	95	73394	57.503	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.000%
Target Compounds						
40) Benzene	6.044	78	32264	7.105	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
Data File : VX027116.D
Acq On : 18 Feb 2022 17:39
Operator : JC/MD
Sample : N1465-02ME
Misc : 4.90g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A5-2-6.5ME

Quant Time: Feb 21 04:04:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

