

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
 Data File : VX027119.D
 Acq On : 18 Feb 2022 18:48
 Operator : JC/MD
 Sample : N1465-09
 Misc : 6.34g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A5-9-9.0

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 02/22/2022
 Supervised By :Semsettin Yesilyurt 02/22/2022

Quant Time: Feb 21 00:42:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	102918	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	167893	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	153166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76795	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	60724	49.152	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.300%		
35) Dibromofluoromethane	5.397	113	48624	48.611	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.220%		
50) Toluene-d8	8.653	98	193624	52.734	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	105.460%		
62) 4-Bromofluorobenzene	11.085	95	75132	58.121	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	116.240%		
Target Compounds						
11) Tert butyl alcohol	3.160	59	49379m	221.248	ug/l	
19) Methyl tert-butyl Ether	3.130	73	24457	6.560	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	6587	1.312	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021822\
Data File : VX027119.D
Acq On : 18 Feb 2022 18:48
Operator : JC/MD
Sample : N1465-09
Misc : 6.34g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A5-9-9.0

Quant Time: Feb 21 00:42:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 02/22/2022
Supervised By :Semsettin Yesilyurt 02/22/2022

