

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021825\
 Data File : VX044986.D
 Acq On : 18 Feb 2025 14:37
 Operator : JC/MD
 Sample : Q1370-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Feb 18 15:29:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	88656	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	178291	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	164456	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	69682	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71090	54.78	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.56%
35) Dibromofluoromethane	5.379	113	60255	51.98	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.96%
50) Toluene-d8	8.647	98	220767	50.36	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.72%
62) 4-Bromofluorobenzene	11.079	95	76472	51.75	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.50%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021825\
Data File : VX044986.D
Acq On : 18 Feb 2025 14:37
Operator : JC/MD
Sample : Q1370-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Feb 18 15:29:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

