

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021825\
 Data File : VX044988.D
 Acq On : 18 Feb 2025 15:23
 Operator : JC/MD
 Sample : PB166736ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB166736ZHE#01

Quant Time: Feb 19 06:13:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	85566	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	171987	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	155329	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66193	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	68557	54.735	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.460%	
35) Dibromofluoromethane	5.379	113	57214	51.166	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.340%	
50) Toluene-d8	8.647	98	213066	50.390	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.780%	
62) 4-Bromofluorobenzene	11.079	95	71961	50.482	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 100.960%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	12536	24.896	ug/l	99
18) Methyl Acetate	2.709	43	2060	1.296	ug/l #	89
20) Methylene Chloride	2.782	84	1428	1.158	ug/l #	85
43) Isopropyl Acetate	6.336	43	97599	33.682	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021825\
Data File : VX044988.D
Acq On : 18 Feb 2025 15:23
Operator : JC/MD
Sample : PB166736ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB166736ZHE#01

Quant Time: Feb 19 06:13:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

