

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021919\
 Data File : VX007608.D
 Acq On : 19 Feb 2019 12:57
 Operator : JC/SP
 Sample : K1518-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CABLE-BRIDGE-FOUNDATION-D-5

Quant Time: Feb 20 06:18:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	431245	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	699770	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	689907	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	320813	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	249858	54.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.18%	
35) Dibromofluoromethane	5.50	113	229742	50.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.92%	
50) Toluene-d8	8.71	98	879830	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
62) 4-Bromofluorobenzene	11.13	95	303197	49.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.32%	
Target Compounds						
16) Acetone	2.45	43	20990	10.011	ug/l	92
18) Methyl Acetate	2.78	43	10926	2.569	ug/l	97
20) Methylene Chloride	2.86	84	14980	3.129	ug/l #	93
43) Isopropyl Acetate	6.46	43	21083	2.152	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021919\
Data File : VX007608.D
Acq On : 19 Feb 2019 12:57
Operator : JC/SP
Sample : K1518-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5

Quant Time: Feb 20 06:18:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

