

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021919\
 Data File : VX007614.D
 Acq On : 19 Feb 2019 15:37
 Operator : JC/SP
 Sample : K1528-02DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2DL

Manual Integrations
 APPROVED

MMDadoda
 2/20/2019 10:34:42 AM

Quant Time: Feb 20 07:08:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	496978	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	773963	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	740649	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	380467	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.07	65	260155	49.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.64%	
35) Dibromofluoromethane	5.50	113	245032	48.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.32%	
50) Toluene-d8	8.71	98	935202	50.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.10%	
62) 4-Bromofluorobenzene	11.13	95	347473	50.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.88%	

Target Compounds

						Qvalue
16) Acetone	2.44	43	5331m	2.206	ug/l	
31) Cyclohexane	5.58	56	153403	20.913	ug/l	94
39) Methylcyclohexane	7.46	83	347964	41.415	ug/l	95
44) Trichloroethene	7.21	130	2089	0.339	ug/l	82
64) Tetrachloroethene	9.34	164	14694	2.105	ug/l	92
67) Ethyl Benzene	10.25	91	28452	1.093	ug/l	98
68) m/p-Xylenes	10.35	106	36231	3.502	ug/l	95
73) Isopropylbenzene	11.02	105	110839	4.287	ug/l	99
78) n-propylbenzene	11.36	91	609047	21.482	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	658273	31.121	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	1688633	78.634	ug/l	99
85) sec-Butylbenzene	11.94	105	67329	2.656	ug/l	98
86) p-Isopropyltoluene	12.06	119	39865	1.742	ug/l	93
95) Naphthalene	13.83	128	47259	1.842	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021919\
Data File : VX007614.D
Acq On : 19 Feb 2019 15:37
Operator : JC/SP
Sample : K1528-02DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2DL

Manual Integrations
APPROVED

MMDadoda
2/20/2019 10:34:42 AM

Quant Time: Feb 20 07:08:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

