

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021924\
 Data File : VX040324.D
 Acq On : 19 Feb 2024 12:44
 Operator : JC/MD
 Sample : P1395-02 10X
 Misc : 4.23g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-3(10-12)

Quant Time: Feb 20 01:12:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	86511	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	151615	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	149084	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80277	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64763	50.320	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 100.640%			
35) Dibromofluoromethane	5.385	113	48564	45.867	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.740%			
50) Toluene-d8	8.653	98	193196	48.779	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.560%			
62) 4-Bromofluorobenzene	11.079	95	80853	50.054	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 100.100%			
Target Compounds						Qvalue
18) Methyl Acetate	2.702	43	6696	4.725	ug/l	95
25) 2-Butanone	4.580	43	1103	1.335	ug/l	96
31) Cyclohexane	5.476	56	2851	1.857	ug/l	96
39) Methylcyclohexane	7.378	83	7109	4.124	ug/l	99
40) Benzene	6.037	78	21872	5.273	ug/l	98
44) Trichloroethene	7.141	130	474	0.414	ug/l	95
52) Toluene	8.750	92	17468928	6429.977	ug/l	77
64) Tetrachloroethene	9.274	164	421	0.403	ug/l #	86
65) Chlorobenzene	10.073	112	1001	0.322	ug/l #	76
67) Ethyl Benzene	10.195	91	2871617	518.205	ug/l	100
68) m/p-Xylenes	10.311	106	3863349	1845.955	ug/l	95
69) o-Xylene	10.646	106	1581026	777.381	ug/l	99
73) Isopropylbenzene	10.963	105	127286	22.512	ug/l	99
78) n-propylbenzene	11.305	91	38235	5.586	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	14596	3.066	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	32458	6.740	ug/l	99
86) p-Isopropyltoluene	12.006	119	11069	2.251	ug/l	94
89) n-Butylbenzene	12.335	91	8392	1.768	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021924\
Data File : VX040324.D
Acq On : 19 Feb 2024 12:44
Operator : JC/MD
Sample : P1395-02 10X
Misc : 4.23g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-3(10-12)

Quant Time: Feb 20 01:12:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

