

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021924\
 Data File : VX040325.D
 Acq On : 19 Feb 2024 13:07
 Operator : JC/MD
 Sample : P1395-03 40X
 Misc : 3.46g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-2(4-6)

Quant Time: Feb 20 01:13:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	81029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	146904	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	140963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74642	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61570	51.075	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 102.160%			
35) Dibromofluoromethane	5.385	113	47299	46.105	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 92.200%			
50) Toluene-d8	8.647	98	181726	47.354	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 94.700%			
62) 4-Bromofluorobenzene	11.079	95	73050	46.674	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 93.340%			
Target Compounds					Qvalue	
39) Methylcyclohexane	7.379	83	17992	10.773	ug/l	93
40) Benzene	6.038	78	52957	13.177	ug/l	99
44) Trichloroethene	7.123	130	32728	29.523	ug/l	97
52) Toluene	8.720	92	2918221	1108.587	ug/l	99
65) Chlorobenzene	10.080	112	259648	88.257	ug/l	98
67) Ethyl Benzene	10.195	91	104189	19.885	ug/l	98
68) m/p-Xylenes	10.299	106	173336	87.594	ug/l	97
69) o-Xylene	10.640	106	44246	23.009	ug/l	96
70) Styrene	10.653	104	29967	9.373	ug/l #	83
73) Isopropylbenzene	10.964	105	5505	1.047	ug/l	100
78) n-propylbenzene	11.305	91	6585	1.035	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	5969	1.349	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	13056	2.916	ug/l	100
86) p-Isopropyltoluene	12.006	119	5562	1.217	ug/l	96
87) 1,3-Dichlorobenzene	11.969	146	2799	1.138	ug/l	89
88) 1,4-Dichlorobenzene	12.043	146	45263	17.374	ug/l	100
89) n-Butylbenzene	12.335	91	6421	1.455	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	14034	5.892	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	2178	1.336	ug/l	93
95) Naphthalene	13.774	128	60012	11.653	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	1813	1.166	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021924\
Data File : VX040325.D
Acq On : 19 Feb 2024 13:07
Operator : JC/MD
Sample : P1395-03 40X
Misc : 3.46g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-2(4-6)

Quant Time: Feb 20 01:13:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

