

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021924\
 Data File : VX040327.D
 Acq On : 19 Feb 2024 13:53
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Feb 20 01:14:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	85334	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	151314	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	144216	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	72267	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64989	51.192	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 102.380%			
35) Dibromofluoromethane	5.385	113	49216	46.575	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.160%			
50) Toluene-d8	8.647	98	185455	46.918	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 93.840%			
62) 4-Bromofluorobenzene	11.079	95	72030	44.681	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 89.360%			
Target Compounds						
52) Toluene	8.720	92	8395	3.096	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021924\
Data File : VX040327.D
Acq On : 19 Feb 2024 13:53
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Feb 20 01:14:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

