

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021924\
 Data File : VX040330.D
 Acq On : 19 Feb 2024 15:03
 Operator : JC/MD
 Sample : P1395-03DL 800X
 Misc : 3.46g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-2(4-6)DL

Quant Time: Feb 20 01:15:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	87816	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	156277	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	145401	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72075	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67471	51.645	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.280%			
35) Dibromofluoromethane	5.385	113	50144	45.947	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.900%			
50) Toluene-d8	8.646	98	187959	46.041	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 92.080%			
62) 4-Bromofluorobenzene	11.079	95	70986	42.635	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 85.260%			
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	2841	1.270	ug/l #	85
44) Trichloroethene	7.135	130	1432	1.214	ug/l	93
52) Toluene	8.720	92	118312	42.249	ug/l	92
65) Chlorobenzene	10.079	112	9929	3.272	ug/l	98
68) m/p-Xylenes	10.299	106	6906	3.383	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021924\
Data File : VX040330.D
Acq On : 19 Feb 2024 15:03
Operator : JC/MD
Sample : P1395-03DL 800X
Misc : 3.46g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-2(4-6)DL

Quant Time: Feb 20 01:15:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

