

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021924\
 Data File : VX040339.D
 Acq On : 19 Feb 2024 18:30
 Operator : JC/MD
 Sample : P1440-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-2

Quant Time: Feb 20 01:20:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74742	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	134473	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128340	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65899	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56792	51.075	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.140%
35) Dibromofluoromethane	5.385	113	43434	46.251	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.500%
50) Toluene-d8	8.647	98	162710	46.319	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.640%
62) 4-Bromofluorobenzene	11.079	95	64504	45.023	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.040%
Target Compounds						
16) Acetone	2.380	43	2165	4.741	ug/l	Qvalue 94
17) Carbon Disulfide	2.514	76	1077	0.535	ug/l #	87
52) Toluene	8.720	92	83117	34.494	ug/l	92
67) Ethyl Benzene	10.195	91	3958	0.830	ug/l	97
68) m/p-Xylenes	10.305	106	1177	0.653	ug/l	99
69) o-Xylene	10.640	106	480	0.274	ug/l	67
73) Isopropylbenzene	10.963	105	9689	2.087	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	1417	0.358	ug/l	92
95) Naphthalene	13.780	128	1116	0.245	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021924\
Data File : VX040339.D
Acq On : 19 Feb 2024 18:30
Operator : JC/MD
Sample : P1440-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-2

Quant Time: Feb 20 01:20:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

