

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021925\
 Data File : VX044998.D
 Acq On : 19 Feb 2025 11:43
 Operator : JC/MD
 Sample : VX0219WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0219WBL01

Quant Time: Feb 20 02:35:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	98956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	201046	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	186227	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	85446	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	79349	54.779	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.560%
35) Dibromofluoromethane	5.373	113	67771	51.847	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.700%
50) Toluene-d8	8.641	98	251891	50.961	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.920%
62) 4-Bromofluorobenzene	11.079	95	89636	53.793	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021925\
Data File : VX044998.D
Acq On : 19 Feb 2025 11:43
Operator : JC/MD
Sample : VX0219WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0219WBL01

Quant Time: Feb 20 02:35:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

