

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021925\
 Data File : VX045004.D
 Acq On : 19 Feb 2025 14:04
 Operator : JC/MD
 Sample : Q1376-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW32DL

Quant Time: Feb 20 02:38:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	87497	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	177846	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	162028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73971	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	71256	55.634	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.260%
35) Dibromofluoromethane	5.379	113	59094	51.106	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.220%
50) Toluene-d8	8.647	98	217606	49.768	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.540%
62) 4-Bromofluorobenzene	11.079	95	78542	53.284	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.560%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	914	1.775	ug/l #	79
17) Carbon Disulfide	2.508	76	823	0.266	ug/l #	87
31) Cyclohexane	5.464	56	27184	13.667	ug/l	98
39) Methylcyclohexane	7.373	83	7301	3.385	ug/l	88
40) Benzene	6.031	78	136868	26.274	ug/l	99
52) Toluene	8.720	92	8739	2.817	ug/l	97
67) Ethyl Benzene	10.195	91	2050	0.334	ug/l #	85
68) m/p-Xylenes	10.299	106	6509	2.873	ug/l	99
69) o-Xylene	10.640	106	765	0.337	ug/l	84
73) Isopropylbenzene	10.963	105	13633	2.289	ug/l	99
78) n-propylbenzene	11.305	91	27520	4.003	ug/l	98
95) Naphthalene	13.780	128	6107	1.144	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021925\
Data File : VX045004.D
Acq On : 19 Feb 2025 14:04
Operator : JC/MD
Sample : Q1376-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW32DL

Quant Time: Feb 20 02:38:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

