

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022020\
 Data File : VX014972.D
 Acq On : 20 Feb 2020 12:14
 Operator : JC/SP
 Sample : L1581-01 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40729

Quant Time: Feb 21 00:50:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	444822	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	701314	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	657200	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	334092	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	256013	51.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.00%	
35) Dibromofluoromethane	5.48	113	214596	49.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.54%	
50) Toluene-d8	8.71	98	836098	50.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.60%	
62) 4-Bromofluorobenzene	11.13	95	303432	51.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.36%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022020\
Data File : VX014972.D
Acq On : 20 Feb 2020 12:14
Operator : JC/SP
Sample : L1581-01 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40729

Quant Time: Feb 21 00:50:42 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

