

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022020\
 Data File : VX014976.D
 Acq On : 20 Feb 2020 13:44
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 21 01:05:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	459132	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	718412	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	697498	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	387506	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	254455	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
35) Dibromofluoromethane	5.49	113	216611	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
50) Toluene-d8	8.71	98	869090	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
62) 4-Bromofluorobenzene	11.13	95	336840	55.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.94%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	11.80	105	6827	0.334	ug/l	97
95) Naphthalene	13.83	128	9834	0.420	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022020\
Data File : VX014976.D
Acq On : 20 Feb 2020 13:44
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 21 01:05:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

