

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022020\
 Data File : VX014978.D
 Acq On : 20 Feb 2020 15:22
 Operator : JC/SP
 Sample : L1605-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HAM-TB021920

Quant Time: Feb 21 01:10:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	469193	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	737927	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	713468	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	387643	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	264104	49.88	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	99.76%	
35) Dibromofluoromethane	5.48	113	222130	48.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.92%	
50) Toluene-d8	8.71	98	880722	50.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.72%	
62) 4-Bromofluorobenzene	11.13	95	340744	54.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.24%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022020\
Data File : VX014978.D
Acq On : 20 Feb 2020 15:22
Operator : JC/SP
Sample : L1605-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HAM-TB021920

Quant Time: Feb 21 01:10:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

