

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022020\
 Data File : VX014980.D
 Acq On : 20 Feb 2020 16:07
 Operator : JC/SP
 Sample : L1603-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HAM-WC-L-02

Quant Time: Feb 21 01:21:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	451809	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	702130	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	681650	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	375890	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	258304	50.66	ug/l	0.00
Spiked Amount			Recovery	=	101.32%	
35) Dibromofluoromethane	5.49	113	212694	49.27	ug/l	0.00
Spiked Amount			Recovery	=	98.54%	
50) Toluene-d8	8.71	98	843566	51.20	ug/l	0.00
Spiked Amount			Recovery	=	102.40%	
62) 4-Bromofluorobenzene	11.13	95	324966	54.75	ug/l	0.00
Spiked Amount			Recovery	=	109.50%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022020\
Data File : VX014980.D
Acq On : 20 Feb 2020 16:07
Operator : JC/SP
Sample : L1603-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HAM-WC-L-02

Quant Time: Feb 21 01:21:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

