

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022023\
 Data File : VX034175.D
 Acq On : 20 Feb 2023 10:02
 Operator : JC/MD
 Sample : VX0220WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0220WBL01

Quant Time: Feb 21 01:03:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	403935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	676001	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	594462	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	265133	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	242205	44.523	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.040%
35) Dibromofluoromethane	5.385	113	212772	48.342	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.680%
50) Toluene-d8	8.647	98	806335	51.025	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.060%
62) 4-Bromofluorobenzene	11.079	95	282569	46.747	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.500%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022023\
Data File : VX034175.D
Acq On : 20 Feb 2023 10:02
Operator : JC/MD
Sample : VX0220WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0220WBL01

Quant Time: Feb 21 01:03:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

