

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022023\
 Data File : VX034187.D
 Acq On : 20 Feb 2023 15:00
 Operator : JC/MD
 Sample : 01593-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-1

Quant Time: Feb 21 01:08:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	347535	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	577693	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	511238	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	220985	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	204745	43.744	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.480%
35) Dibromofluoromethane	5.385	113	182377	48.487	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.980%
50) Toluene-d8	8.647	98	688608	50.991	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.980%
62) 4-Bromofluorobenzene	11.079	95	235712	45.632	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.260%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	16156	6.691	ug/l	93
30) Chloroform	5.099	83	6522	0.885	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022023\
Data File : VX034187.D
Acq On : 20 Feb 2023 15:00
Operator : JC/MD
Sample : 01593-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-1

Quant Time: Feb 21 01:08:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

